

## BEST Interface

Generated by Doxygen 1.8.17



---

<b>1 libbest</b>	<b>1</b>
1.0.1 changelog version . . . . .	1
1.1 Before building . . . . .	1
1.2 Generated files . . . . .	1
1.2.1 Build issues: . . . . .	1
1.3 TEST . . . . .	2
1.4 Changelog . . . . .	2
1.4.1 1.8-19 . . . . .	2
1.4.2 1.8-18 . . . . .	2
1.4.3 1.8-17 . . . . .	2
1.4.4 1.8-16 . . . . .	2
1.4.5 1.8-15 . . . . .	2
1.4.6 1.8-14 . . . . .	2
1.4.7 1.8-13 . . . . .	2
1.4.8 1.8-12 . . . . .	3
1.4.9 1.8-11 . . . . .	3
1.4.10 1.8-10 . . . . .	3
1.4.11 1.8-9 . . . . .	3
1.4.12 1.8-8 . . . . .	3
1.4.13 1.8-7 . . . . .	3
1.4.14 1.8-6 . . . . .	3
1.4.15 1.8-5 . . . . .	3
1.5 Function List . . . . .	4
1.5.1 Data functions . . . . .	4
1.5.1.1 getBuffer . . . . .	4
1.5.1.2 getPosArray_sel . . . . .	4
1.5.1.3 getPosArray . . . . .	4
1.5.1.4 getVoltageArray . . . . .	4
1.5.1.5 getVoltageArray_sel . . . . .	4
1.5.1.6 getCurrentArray . . . . .	4
1.5.1.7 getCurrentArray_sel . . . . .	4
1.5.1.8 getFFT . . . . .	4
1.5.1.9 getPosStats . . . . .	4
1.5.1.10 setDisplayFreq . . . . .	4
1.5.1.11 getDisplayFreq . . . . .	4
1.5.1.12 getBPMStats . . . . .	4
1.5.2 Configuration functions . . . . .	4
1.5.2.1 setOffsetSingle . . . . .	4
1.5.2.2 getOffsetSingle . . . . .	4
1.5.2.3 setOffset . . . . .	4
1.5.2.4 setWindAvg . . . . .	4
1.5.2.5 getWindAvg (added in v1.8-9) . . . . .	4

---

---

1.5.2.6 setPIDparams . . . . .	4
1.5.2.7 setPIDparamSingle . . . . .	4
1.5.2.8 getPIDparamSingle . . . . .	4
1.5.2.9 setPIDoutSel . . . . .	5
1.5.2.10 getPIDoutSel . . . . .	5
1.5.2.11 setCrossBar . . . . .	5
1.5.2.12 getCrossBar (1.8-6) . . . . .	5
1.5.2.13 setBPMorient . . . . .	5
1.5.2.14 getBPMorient . . . . .	5
1.5.2.15 setOutputMux . . . . .	5
1.5.2.16 getOutputMux . . . . .	5
1.5.2.17 setBPMscale . . . . .	5
1.5.2.18 getBPMscale . . . . .	5
1.5.2.19 getBPMscaling_sel . . . . .	5
1.5.2.20 getBPMscaling . . . . .	5
1.5.2.21 setBPMoffset . . . . .	5
1.5.2.22 getBPMoffset (1.8-8) . . . . .	5
1.5.2.23 getBPMselector (1.8-5) . . . . .	5
1.5.2.24 setBPMselector . . . . .	5
1.5.2.25 getROCandROI . . . . .	6
1.5.2.26 getROCenable . . . . .	6
1.5.2.27 setROCenable . . . . .	6
1.5.2.28 setROC . . . . .	6
1.5.2.29 setROClimits . . . . .	6
1.5.2.30 getROClimits (1.8-7) . . . . .	6
1.5.2.31 getROCandROI_sel (TODO) . . . . .	6
1.5.3 Access control functions . . . . .	6
1.5.3.1 getLock . . . . .	6
1.5.3.2 getLockStatus . . . . .	6
1.5.4 PID control functions . . . . .	6
1.5.4.1 getPIDstatus . . . . .	6
1.5.4.2 setFEnable . . . . .	6
1.5.4.3 setCtrlReset . . . . .	6
1.5.4.4 setPIDconf . . . . .	6
1.5.4.5 getPIDconf . . . . .	6
1.5.4.6 setSetpoint . . . . .	7
1.5.4.7 getSetpoint . . . . .	7
1.5.4.8 setSetpoint2 (1.8-8) . . . . .	7
1.5.4.9 getSetpoint2 (1.8-8) . . . . .	7
1.5.4.10 IIRcontrollerSetParam . . . . .	7
1.5.4.11 IIRcontrollerCommitParams . . . . .	7
1.5.4.12 IIRcontrollerGetParam . . . . .	7

---

1.5.5 Function gen control functions . . . . .	7
1.5.5.1 funcGenConfig . . . . .	7
1.5.5.2 funcGenEnable . . . . .	7
1.5.6 TetrAMM functions . . . . .	7
1.5.6.1 tetrammHVSetVoltage . . . . .	7
1.5.6.2 tetrammSetRange . . . . .	7
1.5.6.3 tetrammHVenable . . . . .	7
1.5.7 EnBOX functions . . . . .	7
1.5.7.1 enboxEncoderSelect . . . . .	8
1.5.7.2 enboxMathSelect . . . . .	8
1.5.8 Misc functions . . . . .	8
1.5.8.1 getSystemVersion . . . . .	8
1.5.8.2 getSFPinfo . . . . .	8
<b>2 Module Index</b>	<b>9</b>
2.1 Modules . . . . .	9
<b>3 Class Index</b>	<b>11</b>
3.1 Class List . . . . .	11
<b>4 File Index</b>	<b>13</b>
4.1 File List . . . . .	13
<b>5 Module Documentation</b>	<b>15</b>
5.1 Data functions . . . . .	15
5.1.1 Detailed Description . . . . .	16
5.1.2 Function Documentation . . . . .	16
5.1.2.1 __attribute__() [1/2] . . . . .	16
5.1.2.2 __attribute__() [2/2] . . . . .	16
5.1.2.3 getBPMStats() . . . . .	16
5.1.2.4 getBuffer() . . . . .	16
5.1.2.5 getCurrentArray() . . . . .	16
5.1.2.6 getCurrentArray_sel() . . . . .	17
5.1.2.7 getDisplayFreq() . . . . .	17
5.1.2.8 getFFT() . . . . .	17
5.1.2.9 getPosArray() . . . . .	18
5.1.2.10 getPosArray_sel() . . . . .	18
5.1.2.11 getPosScalar_sel() . . . . .	18
5.1.2.12 getPosStats() . . . . .	19
5.1.2.13 getVoltageArray() . . . . .	19
5.1.2.14 getVoltageArray_sel() . . . . .	19
5.1.2.15 setDisplayFreq() . . . . .	20
5.2 Configuration functions . . . . .	21
5.2.1 Detailed Description . . . . .	22

---

---

5.2.2 Enumeration Type Documentation . . . . .	22
5.2.2.1 best_pid_select . . . . .	22
5.2.3 Function Documentation . . . . .	23
5.2.3.1 getBPMoffset() . . . . .	23
5.2.3.2 getBPMoffsetEnbox() . . . . .	23
5.2.3.3 getBPMorient() . . . . .	23
5.2.3.4 getBPMscale() . . . . .	24
5.2.3.5 getBPMscaling() . . . . .	24
5.2.3.6 getBPMscaling_sel() . . . . .	24
5.2.3.7 getBPMselector() . . . . .	25
5.2.3.8 getCrossBar() . . . . .	25
5.2.3.9 getOffsetSingle() . . . . .	26
5.2.3.10 getOutputMux() . . . . .	26
5.2.3.11 getPIDoutSel() . . . . .	26
5.2.3.12 getPIDparamSingle() . . . . .	27
5.2.3.13 getROCanROI() . . . . .	27
5.2.3.14 getROCenable() . . . . .	28
5.2.3.15 getROClimits() . . . . .	28
5.2.3.16 getWindAvg() . . . . .	28
5.2.3.17 setBPMdevice() . . . . .	29
5.2.3.18 setBPMoffset() . . . . .	29
5.2.3.19 setBPMoffsetEnbox() . . . . .	30
5.2.3.20 setBPMorient() . . . . .	30
5.2.3.21 setBPMscale() . . . . .	30
5.2.3.22 setBPMselector() . . . . .	32
5.2.3.23 setCrossBar() . . . . .	32
5.2.3.24 setEnboxSumDiff() . . . . .	33
5.2.3.25 setEnboxType() . . . . .	33
5.2.3.26 setOffset() . . . . .	33
5.2.3.27 setOffsetSingle() . . . . .	34
5.2.3.28 setOutputMux() . . . . .	34
5.2.3.29 setPIDoutSel() . . . . .	35
5.2.3.30 setPIDparams() . . . . .	35
5.2.3.31 setPIDparamSingle() . . . . .	36
5.2.3.32 setROCenable() . . . . .	36
5.2.3.33 setROClimits() . . . . .	37
5.2.3.34 setWindAvg() . . . . .	37
5.3 Access control functions . . . . .	38
5.3.1 Detailed Description . . . . .	38
5.3.2 Function Documentation . . . . .	38
5.3.2.1 getLock() . . . . .	38
5.3.2.2 getLockStatus() . . . . .	38

---

---

5.4 PID control functions . . . . .	39
5.4.1 Detailed Description . . . . .	39
5.4.2 Function Documentation . . . . .	39
5.4.2.1 getPIDconf() . . . . .	39
5.4.2.2 getPIDstatus() . . . . .	40
5.4.2.3 getSetpoint() . . . . .	40
5.4.2.4 getSetpoint2() . . . . .	40
5.4.2.5 IIRcontrollerCommitParams() . . . . .	41
5.4.2.6 IIRcontrollerGetParam() . . . . .	41
5.4.2.7 IIRcontrollerSetParam() . . . . .	41
5.4.2.8 setCtrlReset() . . . . .	42
5.4.2.9 setFEnable() . . . . .	42
5.4.2.10 setPIDconf() . . . . .	42
5.4.2.11 setSetpoint() . . . . .	43
5.4.2.12 setSetpoint2() . . . . .	43
5.5 Function generator functions . . . . .	44
5.5.1 Detailed Description . . . . .	44
5.5.2 Function Documentation . . . . .	44
5.5.2.1 funcGenConfig() . . . . .	44
5.5.2.2 funcGenEnable() . . . . .	44
5.6 TetrAMM functions . . . . .	45
5.6.1 Detailed Description . . . . .	45
5.6.2 Function Documentation . . . . .	45
5.6.2.1 getTetramms() . . . . .	45
5.6.2.2 getTetrammsNumber() . . . . .	45
5.6.2.3 tetrammHVenable() . . . . .	45
5.6.2.4 tetrammHVSetVoltage() . . . . .	46
5.6.2.5 tetrammSetRange() . . . . .	46
5.7 EnBOX functions . . . . .	47
5.7.1 Detailed Description . . . . .	47
5.7.2 Function Documentation . . . . .	47
5.7.2.1 enboxEncoderSelect() . . . . .	47
5.7.2.2 getEnboxes() . . . . .	47
5.7.2.3 getEnboxesNumber() . . . . .	47
5.8 Misc . . . . .	49
5.8.1 Detailed Description . . . . .	49
5.8.2 Function Documentation . . . . .	49
5.8.2.1 getSFPinfo() . . . . .	49
5.8.2.2 getSystemVersion() . . . . .	49
<b>6 Class Documentation</b>	<b>51</b>
6.1 best_buffer_line Struct Reference . . . . .	51

---

6.1.1 Detailed Description . . . . .	51
6.2 best_stats Struct Reference . . . . .	51
6.2.1 Detailed Description . . . . .	51
<b>7 File Documentation</b> . . . . .	<b>53</b>
7.1 best_c_interface.c File Reference . . . . .	53
7.1.1 Macro Definition Documentation . . . . .	54
7.1.1.1 debug_print . . . . .	55
7.1.1.2 LIB_BEST_DEBUG . . . . .	55
7.1.1.3 VERSION_STRING . . . . .	55
7.1.1.4 VERSION_STRING_HELPER . . . . .	55
7.1.2 Function Documentation . . . . .	55
7.1.2.1 __attribute__() [1/2] . . . . .	55
7.1.2.2 __attribute__() [2/2] . . . . .	55
7.1.2.3 getBestLock() . . . . .	55
7.1.2.4 lock_sigaction() . . . . .	55
7.1.2.5 releaseBestLock() . . . . .	55
7.1.3 Variable Documentation . . . . .	56
7.1.3.1 fd_DMA . . . . .	56
7.1.3.2 fd_Lock . . . . .	56
7.1.3.3 fd_Mbox . . . . .	56
7.1.3.4 in . . . . .	56
7.1.3.5 LIB_BEST_VERSION . . . . .	56
7.1.3.6 loginLevel . . . . .	56
7.1.3.7 out . . . . .	56
7.1.3.8 p . . . . .	56
7.1.3.9 SAMP_FREQ . . . . .	56
7.2 best_c_interface.h File Reference . . . . .	56
7.2.1 Macro Definition Documentation . . . . .	60
7.2.1.1 BEST_FILE_CONFIG_INI . . . . .	60
7.2.1.2 BEST_FILE_DISP_DMA . . . . .	60
7.2.1.3 BEST_FILE_MAILBOX . . . . .	60
7.2.1.4 FFT_LEN . . . . .	60
7.2.1.5 LEN_BBF . . . . .	60
7.2.1.6 LOCK_FILE . . . . .	60
7.2.2 Variable Documentation . . . . .	60
7.2.2.1 LIB_BEST_VERSION . . . . .	60
7.3 best_c_interface_conf.cpp File Reference . . . . .	61
7.3.1 Variable Documentation . . . . .	63
7.3.1.1 fd_DMA . . . . .	63
7.3.1.2 fd_Mbox . . . . .	63
7.3.1.3 in . . . . .	63

7.3.1.4 loginLevel . . . . .	63
7.3.1.5 out . . . . .	63
7.3.1.6 p . . . . .	63
7.3.1.7 SAMP_FREQ . . . . .	63
7.4 best_c_interface_data.cpp File Reference . . . . .	63
7.4.1 Function Documentation . . . . .	64
7.4.1.1 fftw_abs() . . . . .	65
7.4.2 Variable Documentation . . . . .	65
7.4.2.1 fd_DMA . . . . .	65
7.4.2.2 fd_Mbox . . . . .	65
7.4.2.3 in . . . . .	65
7.4.2.4 loginLevel . . . . .	65
7.4.2.5 out . . . . .	65
7.4.2.6 p . . . . .	65
7.4.2.7 SAMP_FREQ . . . . .	65
7.5 best_c_interface_tetramm.cpp File Reference . . . . .	65
7.5.1 Variable Documentation . . . . .	66
7.5.1.1 fd_Mbox . . . . .	66
7.5.1.2 loginLevel . . . . .	66
7.6 README.md File Reference . . . . .	66
<b>Index</b>	<b>67</b>



# **Chapter 1**

## **libbest**

Library to interface with BEST Central Unit FPGA.

### **1.0.1 changelog version**

The last version in `debian/changelog` should be equal to the one in files `version` and in `build_number`. After succesfull compilation the `build_number` and the updated `debian/changelog` should be checked in the version control.

### **1.1 Before building**

- update `.version` and `.buildnumber`
- update `debian` files
  - update `changelog` with `dch -i` accordingly to `version` and `buildnumber`
  - check `debian/control` file (if changes required)
  - update [Readme.md](#) -> Changelog and FunctionList
- commit the changes
- make publish
- doxy `Doxyfile` to update the documentation

### **1.2 Generated files**

Generated `.deb` files are stored in `pkg` folder.

#### **1.2.1 Build issues:**

```
sudo apt install libfftw3-dev
```

## 1.3 TEST

A python script (scanAllGetFunctions.py) can be used to have an idea of who is changing the BEST internal variables

## 1.4 Changelog

### 1.4.1 1.8-19

- Added `getPosScalar` function to get scalar value of PosX, PosY and Intensity
- Modified [getPosStats\(\)](#)

### 1.4.2 1.8-18

- Fixed `setBPMoffsetEnbox` compatible with firmware 1.2.20
- Added `setBPMdevice`, `setEnboxType`, `setEnboxSumDiff` to FPGA for Difference Over Sum Calculations
- Removed `enboxMathSelect`

### 1.4.3 1.8-17

- Added additional channel for `func_gen()` for frequency analysis with PreDAC (before was used the 4-th channel of the predac internally)

### 1.4.4 1.8-16

- Added `setBPMoffsetEnbox()`: to be used to set the BPM offset in [um] only with BEST-ENC

### 1.4.5 1.8-15

- Added [getEnboxes\(\)](#) [getEnboxesNumber\(\)](#) functions

### 1.4.6 1.8-14

- Added `getTetramms` function

### 1.4.7 1.8-13

- Replaced `getBPMscaling_sel` with `getBPMscale` in `getPosArray_sel` and `getFFT`

## 1.4.8 1.8-12

- Compatibility with Best firmware >=1.2.12
- Fixed set/get PidParamSingle

## 1.4.9 1.8-11

- Replaced get getBPMscaling with getBPMscale in getPosStats
- Added function getTetrammsNumber

## 1.4.10 1.8-10

- Removed setROC function
- Added scaling to set/getBPMoffset and set/getROClimits

## 1.4.11 1.8-9

- Added [getWindAvg\(\)](#) -> read PID update frequency from FPGA

## 1.4.12 1.8-8

- Added comments and documentation
- Added [setSetpoint2\(\)](#) -> set new setpoint reading scaling paramenters from FPGA
- Added [getSetpoint2\(\)](#) -> get new setpoint reading scaling paramenters from FPGA
- Added [getBPMoffset\(\)](#) -> read BPM offset from FPGA

## 1.4.13 1.8-7

- Added [getROClimits\(\)](#) -> read ROC limits from FPGA

## 1.4.14 1.8-6

- Added [getCrossBar\(\)](#) -> read Crossbar configuration from FPGA

## 1.4.15 1.8-5

- fixed [getBPMselector\(\)](#) -> read BPM selector configuration from FPGA

## 1.5 Function List

### 1.5.1 Data functions

List of all functions that access real-time data.

---

**1.5.1.1 getBuffer**

---

**1.5.1.2 getPosArray\_sel**

---

**1.5.1.3 getPosArray**

---

**1.5.1.4 getVoltageArray**

---

**1.5.1.5 getVoltageArray\_sel**

---

**1.5.1.6 getCurrentArray**

---

**1.5.1.7 getCurrentArray\_sel**

---

**1.5.1.8 getFFT**

---

**1.5.1.9 getPosStats**

---

**1.5.1.10 setDisplayFreq**

---

**1.5.1.11 getDisplayFreq**

---

**1.5.1.12 getBPMStats**

---

### 1.5.2 Configuration functions

List of all configuration functions.

---

**1.5.2.1 setOffsetSingle**

write single PID offset to FPGA (level 2)

---

**1.5.2.2 getOffsetSingle**

read single PID offset from FPGA (no level)

---

**1.5.2.3 setOffset**

write all 3 PID offsets to FPGA (level 2)

---

**1.5.2.4 setWindAvg**

write average window samples of specific PID to FPGA (level 2)

---

**1.5.2.5 getWindAvg (added in v1.8-9)**

read average window samples of specific PID from FPGA (no level)

---

**1.5.2.6 setPIDparams**

write all the PID parameters of a specific PID to FPGA (level 2)

---

**1.5.2.7 setPIDparamSingle**

write single PID parameter of a specific PID to FPGA (level 2)

---

**1.5.2.8 getPIDparamSingle**

read single PID parameter of a specific PID from FPGA (no level)

**1.5.2.9 setPIDoutSel**

write PID output channel selector to FPGA (level 2) select which PID is assigned to which PreDAC output channel.

**1.5.2.10 getPIDoutSel**

read PID output channel selector from FPGA (no level)

---

**1.5.2.11 setCrossBar**

write TetrAMM crossbar configuration for BPM0 or BPM1 to FPGA (level 2)

**1.5.2.12 getCrossBar (1.8-6)**

read TetrAMM crossbar configuration for BPM0 or BPM1 from FPGA (no level)

---

**1.5.2.13 setBPMorient**

write BPM orientation for BPM0 or BPM1 to FPGA (level 2)

**1.5.2.14 getBPMorient**

read BPM orientation for BPM0 or BPM1 from FPGA (no level)

---

**1.5.2.15 setOutputMux**

write Output Mux configuration to FPGA (level 1, 2)

**1.5.2.16 getOutputMux**

read Output Mux configuration from FPGA (no level)

---

**1.5.2.17 setBPMscale**

write BPM0 or BPM1 scaling parameters to FPGA (level 2)

**1.5.2.18 getBPMscale**

read BPM0 or BPM1 scaling parameters from FPGA (no level)

**1.5.2.19 getBPMscaling\_sel**

read BPM0 or BPM1 scaling parameters from CONFIG\_FILE (no level)

**1.5.2.20 getBPMscaling**

read BPM0 scaling parameters from CONFIG\_FILE (no level)

---

**1.5.2.21 setBPMoffset**

write offsetX and offsetY of BPM0 or BPM1 to FPGA (level 2)

**1.5.2.22 getBPMoffset (1.8-8)**

read offsetX and offsetY of BPM0 or BPM1 from FPGA (no level)

---

**1.5.2.23 getBPMselector (1.8-5)**

read BPM selector configuration from FPGA (no level)

**1.5.2.24 setBPMselector**

write BPM selector configuration to FPGA (no level)!!!

---

**1.5.2.25 getROCandROI**

read RoiX, RoiY, Roil0max, Roil0min and Roc of BPM0 from CONFIG\_FILE (no level)

**1.5.2.26 getROCenable**

read ROC configuration of BPM0 and BPM1 from FPGA (no level)

**1.5.2.27 setROCenable**

write ROC configuration of BPM0 and BPM1 to FPGA (level 2)

**1.5.2.28 setROC**

write min and max ROC levels of specific axis to FPGA (level 2)

**1.5.2.29 setROClimits**

write min and max ROC levels of specific axis to FPGA (no level)

**1.5.2.30 getROClimits (1.8-7)**

read min and max ROC levels of specific axis from FPGA (no level)

**1.5.2.31 getROCandROI\_sel (TODO)**

read RoiX, RoiY, Roil0max, Roil0min and Roc of BPM0 or BPM1 from CONFIG\_FILE (no level)

---

**1.5.3 Access control functions**

List of all control functions.

**1.5.3.1 getLock**

Login function, inputs username and password, output login level

**1.5.3.2 getLockStatus**

returns the actual login level

---

**1.5.4 PID control functions****1.5.4.1 getPIDstatus**

read feedback status from FPGA (no level)

**1.5.4.2 setFBenable**

enable/disable feedback. Write to FPGA (level 1, 2)

**1.5.4.3 setCtrlReset**

reset feedback controller. Write to FPGA (level 1, 2)

**1.5.4.4 setPIDconf**

write PID configuration to FPGA (level 1, 2)

**1.5.4.5 getPIDconf**

read PID configuration from FPGA (no level)

---

**1.5.4.6 setSetpoint**

reads RoiX, RoiY, Roi0max, Roi0min and Roc of BPM0 from CONFIG\_FILE, read BPM-selected scaling parameters (from CONFIG\_FILE), divides the user setpoint by the scaling parameters (from CONFIG\_FILE), sends the unscaled-setpoint to the FPGA (level 1, 2)

**1.5.4.7 getSetpoint**

reads the setpoint from FPGA, reads RoiX, RoiY, Roi0max, Roi0min and Roc of BPM0 from CONFIG\_FILE, read BPM-selected scaling parameters (from CONFIG\_FILE), multiplied the setpoint by the scaling parameters returns the scaled-setpoints to user space

**1.5.4.8 setSetpoint2 (1.8-8)**

read BPM selector configuration from FPGA, read BPM-selected scaling parameters (from FPGA), divides the user setpoint by the scaling parameters, sends the unscaled-setpoint to the FPGA (level 1, 2)

**1.5.4.9 getSetpoint2 (1.8-8)**

reads the setpoint from FPGA, read BPM selector configuration from FPGA, read BPM-selected scaling parameters (from FPGA), multiplied the setpoint by the scaling parameters returns the scaled-setpoints to user space

---

**1.5.4.10 IIRcontrollerSetParam**

write IIR coefficients to FPGA temporarely (level 2)

**1.5.4.11 IIRcontrollerCommitParams**

apply IIR coefficients written to FPGA (level 2)

**1.5.4.12 IIRcontrollerGetParam**

read IIR coefficients from FPGA (no level)

---

**1.5.5 Function gen control functions**

List of all functions to generate sinusoidal signal with PreDAC.

**1.5.5.1 funcGenConfig****1.5.5.2 funcGenEnable****1.5.6 TetrAMM functions**

List of all functions to communicate with TetrAMM.

**1.5.6.1 tetrammHVSetVoltage****1.5.6.2 tetrammSetRange****1.5.6.3 tetrammHVenable****1.5.7 EnBOX functions**

List of all functions to communicate with EnBOX.

---

1.5.7.1 `enboxEncoderSelect`

1.5.7.2 `enboxMathSelect`

## 1.5.8 Misc functions

1.5.8.1 `getSystemVersion`

1.5.8.2 `getSFPinfo`

## Chapter 2

# Module Index

### 2.1 Modules

Here is a list of all modules:

Data functions . . . . .	15
Configuration functions . . . . .	21
Access control functions . . . . .	38
PID control functions . . . . .	39
Function generator functions . . . . .	44
TetrAMM functions . . . . .	45
EnBOX functions . . . . .	47
Misc . . . . .	49



## Chapter 3

# Class Index

### 3.1 Class List

Here are the classes, structs, unions and interfaces with brief descriptions:

<a href="#">best_buffer_line</a> . . . . .	51
<a href="#">best_stats</a> . . . . .	51



## Chapter 4

# File Index

### 4.1 File List

Here is a list of all files with brief descriptions:

<a href="#">best_c_interface.c</a>	.....	53
<a href="#">best_c_interface.h</a>	.....	56
<a href="#">best_c_interface_conf.cpp</a>	.....	61
<a href="#">best_c_interface_data.cpp</a>	.....	63
<a href="#">best_c_interface_tetramm.cpp</a>	.....	65



# Chapter 5

## Module Documentation

### 5.1 Data functions

Functions which read measurement and calculation data from system.

#### Classes

- struct `best_buffer_line`
- struct `best_stats`

#### Functions

- struct `__attribute__((packed)) best_buffer_line`  
• int `getBuffer` (struct `best_buffer_line`[ ], int length)  
*Gets arrays of "samples" from display DMA.*
- int `getPosArray_sel` (int selector, double \*posX, double \*posY, double \*I0, unsigned int length)  
*Gets arrays of positions and intensity from display DMA.*
- int `getPosScalar_sel` (int selector, double \*posX, double \*posY, double \*I0, unsigned int length)  
*Gets arrays of positions and intensity from display DMA.*
- int `getPosArray` (double \*posX, double \*posY, double \*I0, unsigned int length)  
*Same as `getPosArray_sel`, but only for 1st TetrAMM.*
- int `getVoltageArray` (double(\*voltages)[4], unsigned int length)  
*Gets arrays of voltages from display DMA.*
- int `getVoltageArray_sel` (int sel, double(\*voltages)[4], unsigned int length)  
*Gets arrays of voltages one of X PreDAC (the PreDAC is only 1 for the time being)*
- int `getCurrentArray` (double(\*currents)[4], unsigned int length)  
*Gets arrays of currents from display DMA.*
- int `getCurrentArray_sel` (int sel, double(\*currents)[4], unsigned int length)  
*Gets arrays of currents one of two TetrAMMs.*
- int `getFFT` (unsigned int selector, double \*outM, double \*outF, int removeDC)  
*Calculates FFT.*
- int `getPosStats` (int selector, double \*meanX, double \*stdX, double \*stdX\_filt, double \*meanY, double \*stdY, double \*stdY\_filt, double \*meanI0, double \*stdI0, double \*stdI0\_filt)  
*Get position statistics.*
- int `setDisplayFreq` (int freq, uint8\_t use\_filter)  
*Sets display frequency.*
- int `getDisplayFreq` (int \*freq, uint8\_t \*use\_filter)  
*Gets display frequency.*
- struct `__attribute__((__packed__)) best_stats`  
• int `getBPMStats` (`best_stats` \*stats)  
*Gets BPM Statistics.*

### 5.1.1 Detailed Description

Functions which read measurement and calculation data from system.

### 5.1.2 Function Documentation

#### 5.1.2.1 \_\_attribute\_\_() [1/2]

```
struct __attribute__ (
    __packed__ )
```

#### 5.1.2.2 \_\_attribute\_\_() [2/2]

```
struct __attribute__ (
    packed )
```

#### 5.1.2.3 getBPMStats()

```
int getBPMStats (
    best_stats * stats )
```

Gets BPM Statistics.

##### Parameters

<code>stats</code>	in <code>best_stats</code> structure format
--------------------	---

##### Returns

0 on success, -1 on fail

#### 5.1.2.4 getBuffer()

```
int getBuffer (
    struct best_buffer_line [ ],
    int length )
```

Gets arrays of "samples" from display DMA.

##### Parameters

<code>length</code>	number of samples (each sample is 256 bytes big)
---------------------	--

##### Returns

numbers of samples copies (should be equal to length)

#### 5.1.2.5 getCurrentArray()

```
int getCurrentArray (
    double(*) currents[4],
    unsigned int length )
```

Gets arrays of currents from display DMA.

## Parameters

<i>length</i>	Length of the arrays
---------------	----------------------

## Returns

numbers of samples copies (should be equal to length)

**5.1.2.6 getCurrentArray\_sel()**

```
int getCurrentArray_sel (
    int sel,
    double(*) currents[4],
    unsigned int length )
```

Gets arrays of currents one of two TetrAMMs.

## Parameters

<i>sel</i>	TetrAMM selector (0 or 1)
<i>length</i>	Length of the arrays

## Returns

numbers of samples copies (should be equal to length), or EINVAL if wrong arguments

**5.1.2.7 getDisplayFreq()**

```
int getDisplayFreq (
    int * freq,
    uint8_t * use_filter )
```

Gets display frequency.

## Returns

0 on success, -1 on fail

**5.1.2.8 getFFT()**

```
int getFFT (
    unsigned int selector,
    double * outM,
    double * outF,
    int removeDC )
```

Calcuates FFT.

## Parameters

<i>selector</i>	Selects source for FFT ( offset_adc0_ch0 = 24, offset_pos0_posX = 0, offset_pos0_posY = 1, offset_pos0_I0 = 2, offset_pos1_posX = 3, offset_pos1_posY = 4, offset_pos1_I0 = 5, offset_dac0_ch0 = 16, offset_dac0_ch1 = 17, offset_dac0_ch2 = 18, offset_dac0_ch3 = 19 )
<i>outM</i>	magnitude output (size FFT_LEN / 2)
<i>outF</i>	frequency output (size FFT_LEN / 2)
<i>removeDC</i>	removes DC component

### 5.1.2.9 getPosArray()

```
int getPosArray (
    double * posX,
    double * posY,
    double * I0,
    unsigned int length )
```

Same as getPosArray\_sel, but only for 1st TetrAMM.

### 5.1.2.10 getPosArray\_sel()

```
int getPosArray_sel (
    int selector,
    double * posX,
    double * posY,
    double * I0,
    unsigned int length )
```

Gets arrays of postions and intensity from display DMA.

#### Parameters

<i>selector</i>	Selects BPM (= TetrAMM), should be 0 or 1
<i>posX</i>	Pointer to array where position X will be written
<i>posY</i>	Pointer to array where position Y will be written
<i>I0</i>	Pointer to array where Intensity will be written
<i>length</i>	Length of the arrays

#### Returns

numbers of samples copies (should be equal to length)

### 5.1.2.11 getPosScalar\_sel()

```
int getPosScalar_sel (
    int selector,
    double * posX,
    double * posY,
    double * I0,
    unsigned int length )
```

Gets arrays of postions and intensity from display DMA.

#### Parameters

<i>selector</i>	Selects BPM (= TetrAMM), should be 0 or 1
<i>posX</i>	Pointer to array where position X will be written
<i>posY</i>	Pointer to array where position Y will be written
<i>I0</i>	Pointer to array where Intensity will be written
<i>length</i>	Length of the arrays on which to perform the average

## Returns

numbers of samples copies (should be equal to length)

**5.1.2.12 getPosStats()**

```
int getPosStats (
    int selector,
    double * meanX,
    double * stdX,
    double * stdX_filt,
    double * meanY,
    double * stdY,
    double * stdY_filt,
    double * meanI0,
    double * stdI0,
    double * stdI0_filt )
```

Get position statistics.

## Parameters

<i>selector</i>	Selects which BPM
<i>mean</i>	mean value
<i>std</i>	standard deviation value (full bandwidth)
<i>std_filt</i>	standard deviation value (filtered at 100 Hz)

**5.1.2.13 getVoltageArray()**

```
int getVoltageArray (
    double(*) voltages[4],
    unsigned int length )
```

Gets arrays of voltages from display DMA.

## Parameters

<i>length</i>	Length of the arrays
---------------	----------------------

## Returns

numbers of samples copies (should be equal to length)

**5.1.2.14 getVoltageArray\_sel()**

```
int getVoltageArray_sel (
    int sel,
    double(*) voltages[4],
    unsigned int length )
```

Gets arrays of voltages one of X PreDAC (the PreDAC is only 1 for the time being)

## Parameters

<i>sel</i>	PreDAC selector (0, for the time being)
<i>length</i>	Length of the arrays

**Returns**

numbers of samples copies (should be equal to length), or EINVAL if wrong arguments

**5.1.2.15 setDisplayFreq()**

```
int setDisplayFreq (
    int freq,
    uint8_t use_filter )
```

Sets display frequency.

**Parameters**

<i>freq</i>	Frequency in hertz
<i>use_filter</i>	enable filter to prevent aliasing

**Returns**

0 on success, -1 on fail

**Note**

Access level must be user

## 5.2 Configuration functions

Functions which configure system.

### Enumerations

- enum `best_pid_select` { `BEST_PID_SELECT_X` = 0, `BEST_PID_SELECT_Y` = 1, `BEST_PID_SELECT_I0` = 2 }

### Functions

- int `getBPMscaling_sel` (int selector, double \*scaleX, double \*scaleY)  
*Gets BPM scaling parameters for position X and Y in [um]. Read values from CONFIG\_FILE.*
- int `getBPMscaling` (double \*scaleX, double \*scaleY)  
*Gets BPM scaling parameters of 1st BPM for position X and Y in [um]. Read values from CONFIG\_FILE.*
- int `getROCanROI` (double \*roiX, double \*roiY, double \*roi0min, double \*roi0max, double \*roc)  
*Gets RoiX, RoiY, Roi0max, Roi0min and Roc of 1st BPM. Read values from CONFIG\_FILE.*
- int `getROCeable` (bool \*rocX, bool \*rocY, bool \*rocl0, bool \*rocX2, bool \*rocY2, bool \*rocl02)  
*Gets enables for different ROC checks of 1st nad 2nd BPMs. Read values from FPGA.*
- int `setROCeable` (bool rocX, bool rocY, bool rocl0, bool rocX2, bool rocY2, bool rocl02)  
*Sets enables for different ROC checks of 1st nad 2nd BPMs. Writes values to FPGA.*
- int `setOffsetSingle` (enum `best_pid_select` pid\_sel, double offset)  
*Sets output offset in [V] (output=this+PID value, updated even when PID is not running). Write offset to FPGA.*
- int `getOffsetSingle` (enum `best_pid_select` pid\_sel, double \*offset)  
*Gets output offset in [V]. Read offset from FPGA.*
- int `setOffset` (double offsetX, double offsetY, double offsetI0)  
*Sets output offsets in [V] (output=this+PID value, updated even when PID is not running).*
- int `setWindAvg` (enum `best_pid_select` pid\_sel, uint32\_t nr\_samp)  
*Sets number of samples for window averaging.*
- int `getWindAvg` (enum `best_pid_select` pid\_sel, uint32\_t \*nr\_samp)  
*Gets number of samples for window averaging. Read number of samples from FPGA.*
- int `setPIDparams` (char pid\_sel, double Kp, double Ki, double Kd, double emin, double lmax, double Omin, double Omax, double Ogain)  
*Sets PID parameters (legacy).*
- int `setPIDparamSingle` (char pid\_sel, int pid\_par, double param\_value)  
*Sets single PID parameter of specific PID. Write PID parameter to FPGA.*
- int `getPIDparamSingle` (char pid\_sel, int pid\_par, double \*param\_value)  
*Gets PID parameters. Read PID parameters from FPGA.*
- int `setPIDoutSel` (uint8\_t posX, uint8\_t posY, uint8\_t I0)  
*Sets PID output cross switch. Selects which PID is assigned to which PreDAC output channel. Write PID output cross switch to FPGA.*
- int `getPIDoutSel` (uint8\_t \*posX, uint8\_t \*posY, uint8\_t \*I0)  
*Gets PID output cross switch. Read PID output cross switch from FPGA.*
- int `setCrossBar` (uint8\_t sel, uint8\_t ch0, uint8\_t ch1, uint8\_t ch2, uint8\_t ch3)  
*Sets input cross switch.*
- int `setBPMorient` (uint8\_t sel, uint8\_t is90deg)  
*Sets BPM orientation for diffOverSum.*
- int `setOutputMux` (uint8\_t sel)  
*Sets value of output mux.*
- int `getOutputMux` (uint8\_t \*sel)  
*Gets value of output mux.*
- int `setBPMscale` (double scaleX, double scaleY, int bpm)

- int [getBPMscale](#) (int selector, double \*scaleX, double \*scaleY)
 

*Sets BPM scaling parameters for position X and Y in [um]. Write BPM scaling parameters to FPGA.*
- int [setBPMoffset](#) (double offsetX, double offsetY, int bpm)
 

*Gets BPM scaling parameters for position X and Y in [um]. Read BPM scaling parameters from FPGA.*
- int [setBPMoffsetEnbox](#) (double offsetX, double offsetY, int bpm, int encType)
 

*Sets BPM offset in [um] ONLY FOR TETRAMMs. Write BPM offsets to FPGA.*
- int [setBPMdevice](#) (uint8\_t sel, uint8\_t isEnbox)
 

*Sets BPM offset in [um] ONLY FOR Enboxes. Write BPM offsets to FPGA. (>= 1.8-16)*
- int [setEnboxType](#) (uint8\_t sel, uint8\_t encType)
 

*Sets BPM device for diffOverSum. Write to FPGA. (>= 1.8-18).*
- int [setEnboxSumDiff](#) (uint8\_t sel, uint8\_t isSumDiff)
 

*Sets the EnBOX type (relative or absolute) . Write to FPGA. (>= 1.8-18).*
- int [getBPMoffset](#) (double \*offsetX, double \*offsetY, int bpm)
 

*Set diffOverSum calculations (for EnBOX). 0:Normal the positions are calculated from enc1 and enc2. 1:SumDiff the positions are calculated from enc1+enc2 and enc1-enc2. Write to FPGA. (>= 1.8-18).*
- int [getBPMoffsetEnbox](#) (double \*offsetX, double \*offsetY, int bpm, int encType)
 

*Gets BPM offsets in [um] ONLY FOR TETRAMMs. Read BPM offsets from FPGA.*
- int [getBPMorient](#) (uint8\_t sel, uint8\_t \*is90deg)
 

*Gets BPM offsets in [um] ONLY FOR Enboxes. Read BPM offsets from FPGA.*
- int [getBPMselector](#) (int \*posX, int \*posY, int \*l0)
 

*Sets BPM orientation for diffOverSum. Read BPM orientation from FPGA.*
- int [setBPMselector](#) (int posX, int posY, int l0)
 

*Gets BPM sources. Read BPM sources from FPGA.*
- int [setBPMselector](#) (int posX, int posY, int l0)
 

*Sets BPM sources. Write BPM sources to FPGA.*
- int [getCrossBar](#) (uint8\_t sel, uint8\_t \*ch0, uint8\_t \*ch1, uint8\_t \*ch2, uint8\_t \*ch3)
 

*Gets input cross switch. Read input cross switch from FPGA.*
- int [setROCLimits](#) (int sel, double min, double max)
 

*Sets ROC min and max values. ROC should be in [um] for positions and in [A] for intensity. Write min and max ROC levels of specific axis to FPGA.*
- int [getROCLimits](#) (int sel, double \*min, double \*max)
 

*Gets ROC. ROC values are in [um] for positions and in [A] for intensity. Read min and max ROC levels of specific axis from FPGA.*

### 5.2.1 Detailed Description

Functions which configure system.

### 5.2.2 Enumeration Type Documentation

#### 5.2.2.1 best\_pid\_select

```
enum best_pid_select
```

Enumerator

BEST_PID_SELECT_X	
BEST_PID_SELECT_Y	
BEST_PID_SELECT_I0	

### 5.2.3 Function Documentation

#### 5.2.3.1 getBPMoffset()

```
int getBPMoffset (
    double * offsetX,
    double * offsetY,
    int bpm )
```

Gets BPM offsets in [um] ONLY FOR TETRAMMs. Read BPM offsets from FPGA.

##### Parameters

<i>offsetX</i>	in [um]
<i>offsetY</i>	in [um]
<i>bpm</i>	Selects BPM (= TetrAMM), should be 0 or 1

##### Returns

0 on success, negative on fail

#### 5.2.3.2 getBPMoffsetEnbox()

```
int getBPMoffsetEnbox (
    double * offsetX,
    double * offsetY,
    int bpm,
    int encType )
```

Gets BPM offsets in [um] ONLY FOR Enboxes. Read BPM offsets from FPGA.

##### Parameters

<i>offsetX</i>	in [um]
<i>offsetY</i>	in [um]
<i>bpm</i>	Selects BPM (= TetrAMM), should be 0 or 1
<i>encType</i>	Relative (tonic), 1: Absolute (Resolute)

##### Returns

0 on success, negative on fail

#### 5.2.3.3 getBPMorient()

```
int getBPMorient (
    uint8_t sel,
    uint8_t * is90deg )
```

Sets BPM orientation for diffOverSum. Read BPM orientation from FPGA.

##### Parameters

<i>sel</i>	Selects TetrAmm (in case of 2 TetrAmms)
<i>is90deg</i>	(0 = 90deg, 1 = 45deg)

**Returns**

0 on success, negative on fail

**Note**

Access level: all levels

**5.2.3.4 getBPMscale()**

```
int getBPMscale (
    int selector,
    double * scaleX,
    double * scaleY )
```

Gets BPM scaling parameters for position X and Y in [um]. Read BPM scaling parameters from FPGA.

**Parameters**

<i>selector</i>	Selects BPM (= TetrAMM), should be 0 or 1
-----------------	---

**Returns**

0 on success, -1 on fail

**Note**

Access level: all levels

**5.2.3.5 getBPMscaling()**

```
int getBPMscaling (
    double * scaleX,
    double * scaleY )
```

Gets BPM scaling parameters of 1st BPM for position X and Y in [um]. Read values from CONFIG\_FILE.

**Parameters**

<i>scaleX</i>	scaling value for X in [um]
<i>scaleY</i>	scaling value for Y in [um]

**Returns**

0 on success, -1 on fail

**5.2.3.6 getBPMscaling\_sel()**

```
int getBPMscaling_sel (
    int selector,
    double * scaleX,
    double * scaleY )
```

Gets BPM scaling parameters for position X and Y in [um]. Read values from CONFIG\_FILE.

**Parameters**

<i>selector</i>	Selects BPM (= TetrAMM or EnBOX), 0 for 1st BPM, 1 for 2nd BPM
<i>scaleX</i>	pointer to scaling value for X in [um]
<i>scaleY</i>	pointer to scaling value for Y in [um]

**Returns**

0 on success, -1 on fail

**5.2.3.7 getBPMselector()**

```
int getBPMselector (
    int * posX,
    int * posY,
    int * IO )
```

Gets BPM sources. Read BPM sources from FPGA.

**Parameters**

<i>posX</i>	BPM (= TetrAMM) for posX
<i>posY</i>	BPM (= TetrAMM) for posY
<i>IO</i>	BPM (= TetrAMM) for IO

**Returns**

ACK

**Note**

Access level: all levels

**5.2.3.8 getCrossBar()**

```
int getCrossBar (
    uint8_t sel,
    uint8_t * ch0,
    uint8_t * ch1,
    uint8_t * ch2,
    uint8_t * ch3 )
```

Gets input cross switch. Read input cross switch from FPGA.

**Parameters**

<i>sel</i>	Selects TetrAmm (in case of 2 TetrAmms)
<i>ch0</i>	Top / Top-Left
<i>ch1</i>	Right / Top-Right
<i>ch2</i>	Bottom / Bottom-Right
<i>ch3</i>	Left / Bottom-Left

**Returns**

0 on success, negative on fail

**5.2.3.9 getOffsetSingle()**

```
int getOffsetSingle (
    enum best_pid_select pid_sel,
    double * offset )
```

Gets output offset in [V]. Read offset from FPGA.

**Parameters**

<i>pid_sel</i>	PID select enum
<i>offset</i>	offset value in [V]

**Returns**

0 on success, negative on fail

**Note**

Read from FPGA

Access level: all levels

**5.2.3.10 getOutputMux()**

```
int getOutputMux (
    uint8_t * sel )
```

Gets value of output mux.

**Parameters**

<i>sel</i>	Selects between FPGA (value 1) and software (value 0)
------------	---

**Returns**

0 on success, negative on fail

**Note**

Access level: all levels

**5.2.3.11 getPIDoutSel()**

```
int getPIDoutSel (
    uint8_t * posX,
    uint8_t * posY,
    uint8_t * IO )
```

Gets PID output cross switch. Read PID output cross switch from FPGA.

**Parameters**

<i>posX</i>	DAC channel for posX
-------------	----------------------

**Parameters**

<i>posY</i>	DAC channel for posY
<i>I0</i>	DAC channel for I0

**Returns**

0 on success, negative on fail

**Note**

Access level: all levels

**5.2.3.12 getPIDparamSingle()**

```
int getPIDparamSingle (
    char pid_sel,
    int pid_par,
    double * param_value )
```

Gets PID parameters. Read PID parameters from FPGA.

**Parameters**

<i>pid_sel</i>	(0 = X, 1 = Y, 2 = I0)
<i>pid_par</i>	(0=Kp, 1=Ki, 2=Kd, 3=emin, 4=Imax, 5=Omin, 6=Omax, 7=Ogain)
<i>param_value</i>	

**Returns**

0 on success, negative on fail

**Note**

Access level: all levels

**5.2.3.13 getROCandROI()**

```
int getROCandROI (
    double * roiX,
    double * roiY,
    double * roiI0min,
    double * roiI0max,
    double * roc )
```

Gets RoiX, RoiY, RoiI0max, RoiI0min and Roc of 1st BPM. Read values from CONFIG\_FILE.

**Parameters**

<i>roiX</i>	double parameter in [um]
<i>roiY</i>	double parameter in [um]
<i>roiI0min</i>	double parameter in [A]
<i>roiI0max</i>	double parameter in [A]
<i>roc</i>	double parameter in percentage

**Note**

Access level: all levels

**5.2.3.14 getROCenable()**

```
int getROCenable (
    bool * rocX,
    bool * rocY,
    bool * rocI0,
    bool * rocX2,
    bool * rocY2,
    bool * rocI02 )
```

Gets enables for different ROC checks of 1st nad 2nd BPMs. Read values from FPGA.

**Parameters**

<i>rocX</i>	enable rocX
<i>rocY</i>	enable rocY
<i>rocI0</i>	enable rocI0
<i>rocX2</i>	enable rocX2
<i>rocY2</i>	enable rocY2
<i>rocI02</i>	enable rocI02

**Returns**

0 on success, -1 on fail

**Note**

Access level: all levels

**5.2.3.15 getROClimits()**

```
int getROClimits (
    int sel,
    double * min,
    double * max )
```

Gets ROC. ROC values are in [um] for positions and in [A] for intensity. Read min and max ROC levels of specific axis from FPGA.

**Parameters**

<i>sel</i>	(0 = X, 1 = Y, 2 = I0, 0 = X2, 1 = Y2, 2 = I02)
<i>min</i>	
<i>max</i>	

**Returns**

0 on success, negative on fail

**5.2.3.16 getWindAvg()**

```
int getWindAvg (
```

```
enum best_pid_select pid_sel,
    uint32_t * nr_samp )
```

Gets number of samples for window averaging. Read number of samples from FPGA.

#### Parameters

<i>sel</i>	(0 = X, 1 = Y, 2 = I0)
<i>nr_samp</i>	Number of samples

#### Returns

0 on success, negative on fail

### 5.2.3.17 setBPMdevice()

```
int setBPMdevice (
    uint8_t sel,
    uint8_t isEnbox )
```

Sets BPM device for diffOverSum. Write to FPGA. (>= 1.8-18).

#### Parameters

<i>sel</i>	Selects the BPM
<i>isEnbox</i>	

#### Returns

0 on success, negative on fail

#### Note

Access level must be admin

### 5.2.3.18 setBPMoffset()

```
int setBPMoffset (
    double offsetX,
    double offsetY,
    int bpm )
```

Sets BPM offset in [um] ONLY FOR TETRAMMs. Write BPM offsets to FPGA.

#### Parameters

<i>offsetX</i>	in [um]
<i>offsetY</i>	in [um]
<i>bpm</i>	Selects BPM (= TetrAMM), should be 0 or 1

#### Returns

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.19 setBPMoffsetEnbox()**

```
int setBPMoffsetEnbox (
    double offsetX,
    double offsetY,
    int bpm,
    int encType )
```

Sets BPM offset in [um] ONLY FOR Enboxes. Write BPM offsets to FPGA. (>= 1.8-16)

**Parameters**

<i>offsetX</i>	in [um]
<i>offsetY</i>	in [um]
<i>bpm</i>	Selects BPM (= Enbox), should be 0 or 1
<i>encType,0</i>	Relative (tonic), 1: Absolute (Resolute)

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.20 setBPMorient()**

```
int setBPMorient (
    uint8_t sel,
    uint8_t is90deg )
```

Sets BPM orientation for diffOverSum.

**Parameters**

<i>sel</i>	Selects TetrAmm (in case of 2 TetrAmms)
<i>is90deg</i>	

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.21 setBPMscale()**

```
int setBPMscale (
    double scaleX,
    double scaleY,
    int bpm )
```

Sets BPM scaling parameters for position X and Y in [um]. Write BPM scaling parameters to FPGA.

**Parameters**

<i>scaleX</i>	in [um]
<i>scaleY</i>	in [um]
<i>bpm</i>	Selects BPM (= TetrAMM), should be 0 or 1

**Returns**

0 on success, -1 on fail

**Note**

Access level must be admin

**5.2.3.22 setBPMselector()**

```
int setBPMselector (
    int posX,
    int posY,
    int IO )
```

Sets BPM sources. Write BPM sources to FPGA.

**Parameters**

<i>posX</i>	BPM (= TetrAMM) for posX
<i>posY</i>	BPM (= TetrAMM) for posY
<i>IO</i>	BPM (= TetrAMM) for IO

**Returns**

ACK

**Note**

Access level: all levels

**5.2.3.23 setCrossBar()**

```
int setCrossBar (
    uint8_t sel,
    uint8_t ch0,
    uint8_t ch1,
    uint8_t ch2,
    uint8_t ch3 )
```

Sets input cross switch.

**Parameters**

<i>sel</i>	Selects TetrAmm (in case of 2 TetrAmms)
<i>ch0</i>	
<i>ch1</i>	
<i>ch2</i>	
<i>ch3</i>	

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.24 setEnboxSumDiff()**

```
int setEnboxSumDiff (
    uint8_t sel,
    uint8_t isSumDiff )
```

Set diffOverSum calculations (for EnBOX). 0:Normal the positions are calculated from enc1 and enc2. 1:SumDiff the positions are calculated from enc1+enc2 and enc1-enc2. Write to FPGA. (>= 1.8-18).

**Parameters**

<i>sel</i>	Selects the BPM
<i>encType,0</i>	Normal (enc1, enc2), 1: SumDiff (enc1+enc2, enc1-enc2)

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.25 setEnboxType()**

```
int setEnboxType (
    uint8_t sel,
    uint8_t encType )
```

Sets the EnBOX type (relative or absolute) . Write to FPGA. (>= 1.8-18).

**Parameters**

<i>sel</i>	Selects the BPM
<i>encType,0</i>	Relative (tonic), 1: Absolute (Resolute)

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.26 setOffset()**

```
int setOffset (
    double offsetX,
```

```
    double offsetY,
    double offsetIO )
```

Sets output offsets in [V] (output=this+PID value, updated even when PID is not running).

#### Parameters

<i>offsetX</i>	in [V]
<i>offsetY</i>	in [V]
<i>offsetI0</i>	in [V]

#### Returns

0 on success, negative on fail

#### Note

Access level must be admin

### 5.2.3.27 setOffsetSingle()

```
int setOffsetSingle (
    enum best_pid_select pid_sel,
    double offset )
```

Sets output offset in [V] (output=this+PID value, updated even when PID is not running). Write offset to FPGA.

#### Parameters

<i>pid_sel</i>	PID select enum
<i>offset</i>	offset value in [V]

#### Returns

0 on success, negative on fail

#### Note

Write to FPGA

Access level must be admin

### 5.2.3.28 setOutputMux()

```
int setOutputMux (
    uint8_t sel )
```

Sets value of output mux.

#### Parameters

<i>sel</i>	Selects between FPGA (value 1) and software (value 0)
------------	---

**Returns**

0 on success, negative on fail

**Note**

Access level must be user or admin

**5.2.3.29 setPIDoutSel()**

```
int setPIDoutSel (
    uint8_t posX,
    uint8_t posY,
    uint8_t IO )
```

Sets PID output cross switch. Selects which PID is assigned to which PreDAC output channel. Write PID output cross switch to FPGA.

**Parameters**

<i>posX</i>	DAC channel for posX
<i>posY</i>	DAC channel for posY
<i>IO</i>	DAC channel for IO

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.30 setPIDparams()**

```
int setPIDparams (
    char pid_sel,
    double Kp,
    double Ki,
    double Kd,
    double emin,
    double Imax,
    double Omin,
    double Omax,
    double Ogain )
```

Sets PID parameters (legacy).

**Parameters**

<i>pid_sel</i>	(0 = X, 1 = Y, 2 = IO)
<i>Kp</i>	See BEST User's Manual for detailed explanation
<i>Ki</i>	See BEST User's Manual for detailed explanation
<i>Kd</i>	See BEST User's Manual for detailed explanation
<i>emin</i>	See BEST User's Manual for detailed explanation
<i>Imax</i>	See BEST User's Manual for detailed explanation
<i>Omin</i>	See BEST User's Manual for detailed explanation
<i>Omax</i>	See BEST User's Manual for detailed explanation
<i>Ogain</i>	See BEST User's Manual for detailed explanation

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.31 setPIDparamSingle()**

```
int setPIDparamSingle (
    char pid_sel,
    int pid_par,
    double param_value )
```

Sets single PID parameter of specific PID. Write PID parameter to FPGA.

**Parameters**

<i>pid_sel</i>	(0 = X, 1 = Y, 2 = I0)
<i>pid_par</i>	(0=Kp, 1=Ki, 2=Kd, 3=emin, 4=Imax, 5=Omin, 6=Omax, 7=Ogain)
<i>param_value</i>	

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

**5.2.3.32 setROCenable()**

```
int setROCenable (
    bool rocX,
    bool rocY,
    bool rocI0,
    bool rocX2,
    bool rocY2,
    bool rocI02 )
```

Sets enables for different ROC checks of 1st nad 2nd BPMs. Writes values to FPGA.

**Parameters**

<i>rocX</i>	enable rocX
<i>rocY</i>	enable rocY
<i>rocI0</i>	enable rocI0
<i>rocX2</i>	enable rocX2
<i>rocY2</i>	enable rocY2
<i>rocI02</i>	enable rocI02

**Returns**

0 on success, -1 on fail

**Note**

Access level must be admin

**5.2.3.33 setROCLimits()**

```
int setROCLimits (
    int sel,
    double min,
    double max )
```

Sets ROC min and max values. ROC should be in [um] for positions and in [A] for intensity. Write min and max ROC levels of specific axis to FPGA.

**Parameters**

<i>sel</i>	(0 = X, 1 = Y, 2 = I0, 3 = X2, 4 = Y2, 5 = I02)
<i>min</i>	
<i>max</i>	

**Returns**

0 on success, negative on fail

**Note**

Access level: access level user or admin

**5.2.3.34 setWindAvg()**

```
int setWindAvg (
    enum best_pid_select pid_sel,
    uint32_t nr_samp )
```

Sets number of samples for window averaging.

**Parameters**

<i>sel</i>	(0 = X, 1 = Y, 2 = I0)
<i>nr_samp</i>	Number of samples

**Returns**

0 on success, negative on fail

**Note**

Access level must be admin

## 5.3 Access control functions

Functions which allow to elevate user access level.

### Functions

- int [getLock](#) (const char \*usr, const char \*pass)  
*Acquires lock on /var/lock/best\_lock.*
- int [getLockStatus](#) ()  
*Gets login level, which can be changed with call on [getLock\(\)](#).*

### 5.3.1 Detailed Description

Functions which allow to elevate user access level.

### 5.3.2 Function Documentation

#### 5.3.2.1 [getLock\(\)](#)

```
int getLock (
    const char * usr,
    const char * pass )
```

Acquires lock on /var/lock/best\_lock.

##### Parameters

<i>usr</i>	Username
<i>pass</i>	Password

##### Returns

login level (0=cruise, 1=user, 2=admin)

##### Note

Access level: all levels

#### 5.3.2.2 [getLockStatus\(\)](#)

```
int getLockStatus ( )
```

Gets login level, which can be changed with call on [getLock\(\)](#).

##### Returns

login level (0=cruise, 1=user, 2=admin)

##### Note

Access level: all levels

## 5.4 PID control functions

Functions which control PID behaviour.

### Functions

- `unsigned int getPIDstatus (void)`  
*Gets PID status.*
- `int setFEnable (int enable)`  
*Enables or disables feedback.*
- `int setCtrlReset ()`  
*Resets internal states of controller.*
- `int setPIDconf (char conf)`  
*Sets PID controller configuration. Write PID controller configuration to FPGA.*
- `int getPIDconf (char *conf)`  
*Sets PID controller configuration. Read PID configuration from FPGA.*
- `int setSetpoint (char sel, double setpt)`  
*Sets new setpoint in [um]. Write setpoint to FPGA. Before writing to FPGA the setpoint is divided by scaling parameter of 1st BPM. Scaling parameters are read from CONFIG\_FILE.*
- `int getSetpoint (char sel, double *setpt)`  
*Gets setpoint in [um]. Read setpoint from FPGA. After reading from FPGA the setpoint is multiplied by scaling parameter. Scaling parameters are read from CONFIG\_FILE.*
- `int setSetpoint2 (char sel, double setpt)`  
*Sets new setpoint (without using config file). Write setpoint to FPGA. Before writing to FPGA the setpoint is divided by scaling parameter of 1st BPM. Scaling parameters are read from FPGA.*
- `int getSetpoint2 (char sel, double *setpt)`  
*Gets new setpoint (without using config file). Read setpoint from FPGA. After reading from FPGA the setpoint is multiplied by scaling parameter. Scaling parameters are read from FPGA.*
- `int IIRcontrollerSetParam (char ctrlr_sel, char param_is_a, char param_sel, double param)`  
*Sets IIR controller parameter.*
- `int IIRcontrollerCommitParams (char ctrlr_sel)`  
*Commits IIR controller parameters.*
- `int IIRcontrollerGetParam (char ctrlr_sel, char param_is_a, char param_sel, double *param)`  
*Reads IIR controller parameter.*

### 5.4.1 Detailed Description

Functions which control PID behaviour.

### 5.4.2 Function Documentation

#### 5.4.2.1 `getPIDconf()`

```
int getPIDconf (
    char * conf )
```

Sets PID controller configuration. Read PID configuration from FPGA.

Parameters

<code>conf</code>	return PID conf
-------------------	-----------------

**Returns**

0 on success, negative on fail

**5.4.2.2 getPIDstatus()**

```
unsigned int getPIDstatus (
    void )
```

Gets PID status.

**Returns**

stoppedByUser = 0, stoppedByROC = 1, paused = 2, running = 3

**Note**

Access level: all levels

**5.4.2.3 getSetpoint()**

```
int getSetpoint (
    char sel,
    double * setpt )
```

Gets setpoint in [um]. Read setpoint from FPGA. After reading from FPGA the setpoint is multiplied by scaling parameter. Scaling parameters are read from CONFIG\_FILE.

**Parameters**

<i>sel</i>	Selects which PID ( X = 0, Y = 1, I0 = 2 )
<i>setpt</i>	Setpoint in [um]

**Returns**

0 on success, negative on fail

**Note**

Access level: all levels

**5.4.2.4 getSetpoint2()**

```
int getSetpoint2 (
    char sel,
    double * setpt )
```

Gets new setpoint (without using config file). Read setpoint from FPGA. After reading from FPGA the setpoint is multiplied by scaling parameter. Scaling parameters are read from FPGA.

**Parameters**

<i>sel</i>	Selects which PID ( X = 0, Y = 1, I0 = 2 )
<i>setpt</i>	Setpoint in [um]

**Returns**

0 on success, negative on fail

**5.4.2.5 IIRcontrollerCommitParams()**

```
int IIRcontrollerCommitParams (
    char ctrlr_sel )
```

Commits IIR controller parameters.

**Parameters**

<i>ctrlr_sel</i>	Selects controller ( X = 0, Y = 1, IO = 2 )
------------------	---

**Returns**

0 on success, negative on fail

**5.4.2.6 IIRcontrollerGetParam()**

```
int IIRcontrollerGetParam (
    char ctrlr_sel,
    char param_is_a,
    char param_sel,
    double * param )
```

Reads IIR controller parameter.

**Parameters**

<i>ctrlr_sel</i>	Selects controller ( X = 0, Y = 1, IO = 2 )
<i>param_is_a</i>	Selects if param read is a (when 1) or b (when 0)
<i>param_sel</i>	Selects parameter (0-9)
<i>param</i>	IIR coefficient

**Returns**

0 on success, negative on fail

This function returns the parameter from read-back storage (the parameters which are actually being used by IIR controller).

**5.4.2.7 IIRcontrollerSetParam()**

```
int IIRcontrollerSetParam (
    char ctrlr_sel,
    char param_is_a,
    char param_sel,
    double param )
```

Sets IIR controller parameter.

**Parameters**

<i>ctrlr_sel</i>	Selects controller ( X = 0, Y = 1, IO = 2 )
<i>param_is_a</i>	Selects if param written is a (when 1) or b (when 0)

**Parameters**

<i>param_sel</i>	Selects parameter (0-9)
<i>param</i>	IIR coefficient

**Returns**

0 on success, negative on fail

This function stores the parameter in the temporary storage. After all parameters have been set use IIRcontroller←CommitParams function to download parameters to controller.

**5.4.2.8 setCtrlReset()**

```
int setCtrlReset ( )
```

Resets internal states of controller.

**Returns**

0 on success, negative on fail (e.g. authentication)

**Note**

Access level must be user or admin

**5.4.2.9 setFBenable()**

```
int setFBenable (
    int enable )
```

Enables or disables feedback.

**Returns**

0 on success, negative on fail (e.g. authentication)

**Note**

Access level must be user or admin

**5.4.2.10 setPIDconf()**

```
int setPIDconf (
    char conf )
```

Sets PID controller configuration. Write PID controller configuration to FPGA.

**Parameters**

<i>conf</i>	PID conf (X = 0, XY = 1, XI0 = 2, Y = 3, YI0 = 4, XYI0 = 5, I0 = 6)
-------------	---

**Returns**

0 on success, negative on fail (e.g. authentication)

**Note**

Access level must be user or admin

#### 5.4.2.11 setSetpoint()

```
int setSetpoint (
    char sel,
    double setpt )
```

Sets new setpoint in [um]. Write setpoint to FPGA. Before writing to FPGA the setpoint is divided by scaling parameter of 1st BPM. Scaling parameters are read from CONFIG\_FILE.

##### Parameters

<i>sel</i>	Selects which PID ( X = 0, Y = 1, IO = 2 )
<i>setpt</i>	Setpoint in [um]

##### Returns

0 on success, negative on fail (e.g. authentication)

##### Note

Access level must be user or admin

#### 5.4.2.12 setSetpoint2()

```
int setSetpoint2 (
    char sel,
    double setpt )
```

Sets new setpoint (without using config file). Write setpoint to FPGA. Before writing to FPGA the setpoint is divided by scaling parameter of 1st BPM. Scaling parameters are read from FPGA.

##### Parameters

<i>sel</i>	Selects which PID ( X = 0, Y = 1, IO = 2 )
<i>setpt</i>	Setpoint in [um]

##### Returns

0 on success, negative on fail (e.g. authentication)

##### Note

Access level must be user or admin

## 5.5 Function generator functions

### Functions

- int **funcGenConfig** (float freq, float ampl[4], float offset[4])  
*Configures function generator.*
- int **funcGenEnable** (char enable)  
*Enables or disables function generator.*

#### 5.5.1 Detailed Description

#### 5.5.2 Function Documentation

##### 5.5.2.1 funcGenConfig()

```
int funcGenConfig (
    float freq,
    float ampl[4],
    float offset[4] )
```

Configures function generator.

#### Parameters

<i>freq</i>	Frequency (in Hz)
<i>ampl</i>	Amplitude of sine wave
<i>offset</i>	Offset of sine wave

#### Returns

0 on success, negative on fail

configuration is committed when the func gen is enabled. if func gen is already running, just call funcGenEnable function again

##### 5.5.2.2 funcGenEnable()

```
int funcGenEnable (
    char enable )
```

Enables or disables function generator.

#### Parameters

<i>enable</i>	0 to disable, 1 to enable
---------------	---------------------------

## 5.6 TetrAMM functions

### Functions

- int `tetrammSetRange` (int range, unsigned char selector)  
*Sets front-end range.*
- int `tetrammHVSetVoltage` (float voltage, unsigned char sel)  
*Sets TetrAMM HV voltage.*
- int `tetrammHVenable` (int enable, unsigned char sel)  
*Enables or disables TetrAMM HV module.*
- unsigned int `getTetrammsNumber` ()  
*Gets number of TetrAMMs.*
- int `getTetramms` (int \*pos\_A, int \*pos\_B)  
*Gets TetrAMMs. (>= 1.8-14).*

### 5.6.1 Detailed Description

### 5.6.2 Function Documentation

#### 5.6.2.1 `getTetramms()`

```
int getTetramms (
    int * pos_A,
    int * pos_B )
```

Gets TetrAMMs. (>= 1.8-14).

##### Parameters

<code>pos_A</code>	on SFP A, 0 = No TetrAMM connected, 1 = TetrAMM connected
<code>pos_A</code>	on SFP B, 0 = No TetrAMM connected, 1 = TetrAMM connected

##### Returns

0 on success, negative on fail

#### 5.6.2.2 `getTetrammsNumber()`

```
unsigned int getTetrammsNumber ( )
```

Gets number of TetrAMMs.

##### Returns

TetrAMMs number

#### 5.6.2.3 `tetrammHVenable()`

```
int tetrammHVenable (
    int enable,
    unsigned char sel )
```

Enables or disables TetrAMM HV module.

**Parameters**

<i>enable</i>	1 to enable, 0 to disable
<i>sel</i>	0 = TetrAMM on SFP A, 1 = TetrAMM on SFP B

**Note**

Access level must be admin

**5.6.2.4 tetrammHVSetVoltage()**

```
int tetrammHVSetVoltage (
    float voltage,
    unsigned char sel )
```

Sets TetrAMM HV voltage.

**Parameters**

<i>voltage</i>	voltage in volts
<i>sel</i>	0 = TetrAMM on SFP A, 1 = TetrAMM on SFP B

**Note**

Access level must be admin

**5.6.2.5 tetrammSetRange()**

```
int tetrammSetRange (
    int range,
    unsigned char selector )
```

Sets front-end range.

**Parameters**

<i>range</i>	0 = RNG0, 1 = RNG1
<i>selector</i>	0 = TetrAMM on SFP A, 1 = TetrAMM on SFP B

**Note**

Access level must be admin

## 5.7 EnBOX functions

### Functions

- int [enboxEncoderSelect](#) (int enc\_type, unsigned char selector)
- unsigned int [getEnboxesNumber](#) ()  
*Gets number of Enbox. (>= 1.8-15).*
- int [getEnboxes](#) (int \*pos\_A, int \*pos\_B)  
*Gets Enbox. (>= 1.8-15).*

#### 5.7.1 Detailed Description

#### 5.7.2 Function Documentation

##### 5.7.2.1 enboxEncoderSelect()

```
int enboxEncoderSelect (
    int enc_type,
    unsigned char selector )
```

#### Parameters

<i>enc_type</i>	0 = Tonic, 1 = Absolute
<i>selector</i>	0 = EnBOX on SFP A, 1 = EnBOX on SFP B

#### Note

Access level must be admin

##### 5.7.2.2 getEnboxes()

```
int getEnboxes (
    int * pos_A,
    int * pos_B )
Gets Enbox. (>= 1.8-15).
```

#### Parameters

<i>pos_A</i>	on SFP A, 0 = No Enbox connected, 1 = Enbox connected
<i>pos_A</i>	on SFP B, 0 = No Enbox connected, 1 = Enbox connected

#### Returns

0 on success, negative on fail

##### 5.7.2.3 getEnboxesNumber()

```
unsigned int getEnboxesNumber ( )
Gets number of Enbox. (>= 1.8-15).
```

**Returns**

Enboxes number

## 5.8 Misc

### Functions

- int [getSFPinfo](#) (int selector, uint16\_t \*id, uint32\_t \*timestamp, uint32\_t \*fw\_ver, uint32\_t \*ser\_nr)  
*Gets SFP info.*
- int [getSystemVersion](#) (uint32\_t \*hwVer, uint32\_t \*swVer, uint32\_t \*ctrlVer)  
*Gets System HW, SW and PID versions.*

#### 5.8.1 Detailed Description

#### 5.8.2 Function Documentation

##### 5.8.2.1 getSFPinfo()

```
int getSFPinfo (
    int selector,
    uint16_t * id,
    uint32_t * timestamp,
    uint32_t * fw_ver,
    uint32_t * ser_nr )
```

Gets SFP info.

#### Parameters

<i>selector</i>	Selects which SFP port (A=0, B=1, ...)
<i>id</i>	
<i>timestamp</i>	
<i>fw_ver</i>	
<i>ser_nr</i>	

#### Returns

0 on success, -1 on fail

#### Note

Access level: all levels

##### 5.8.2.2 getSystemVersion()

```
int getSystemVersion (
    uint32_t * hwVer,
    uint32_t * swVer,
    uint32_t * ctrlVer )
```

Gets System HW, SW and PID versions.

#### Parameters

<i>hwVer</i>	hardware version
<i>swVer</i>	software version (main application)
<i>ctrlVer</i>	controller version (pid version)

**Returns**

0 on success, -1 on fail

**Note**

Access level: all levels

# Chapter 6

## Class Documentation

### 6.1 best\_buffer\_line Struct Reference

```
#include <best_c_interface.h>
```

#### 6.1.1 Detailed Description

best DMA buffer struct

The documentation for this struct was generated from the following file:

- [best\\_c\\_interface.h](#)

### 6.2 best\_stats Struct Reference

```
#include <best_c_interface.h>
```

#### 6.2.1 Detailed Description

best statistics data structure

The documentation for this struct was generated from the following file:

- [best\\_c\\_interface.h](#)



# Chapter 7

## File Documentation

### 7.1 best\_c\_interface.c File Reference

```
#include "best_c_interface.h"
#include <stdlib.h>
#include <time.h>
#include <stdint.h>
#include <string.h>
#include <errno.h>
#include <unistd.h>
#include <fcntl.h>
#include <sys/stat.h>
#include <sys/ioctl.h>
#include "pcie_driver/BEST_PCIE.h"
#include "pcie_mailbox/mailbox_comm_defs.h"
#include <iostream>
#include "inlh/cpp/INIReader.h"
#include <complex>
#include <fftw3.h>
#include <signal.h>
```

#### Macros

- #define VERSION\_STRING\_HELPER(X) #X
- #define VERSION\_STRING(X) VERSION\_STRING\_HELPER(X)
- #define LIB\_BEST\_DEBUG 0
- #define debug\_print(fmt, ...) do { if (LIB\_BEST\_DEBUG) fprintf(stderr, "[libbest] " fmt, \_\_VA\_ARGS\_\_); } while (0)

#### Functions

- int **getBestLock** (int fd)  
*!!! All sets should check loginLevel*
- int **releaseBestLock** (int fd)
- void **lock\_sigaction** (int sig, siginfo\_t \*siginfo, void \*context)
- void **\_\_attribute\_\_ ((constructor))**
- void **\_\_attribute\_\_ ((destructor))**
- int **getLock** (const char \*usr, const char \*pass)  
*Acquires lock on /var/lock/best\_lock.*
- int **getLockStatus** ()  
*Gets login level, which can be changed with call on getLock().*
- int **setFEnable** (int enable)

- `int setCtrlReset ()`  
*Enables or disables feedback.*
- `int setPIDconf (char conf)`  
*Resets internal states of controller.*
- `int getPIDconf (char *conf)`  
*Sets PID controller configuration. Write PID controller configuration to FPGA.*
- `int setSetpoint (char sel, double setpt)`  
*Sets PID controller configuration. Read PID configuration from FPGA.*
- `int setSetpoint2 (char sel, double setpt)`  
*Sets new setpoint in [um]. Write setpoint to FPGA. Before writing to FPGA the setpoint is divided by scaling parameter of 1st BPM. Scaling parameters are read from CONFIG\_FILE.*
- `int getSetpoint (char sel, double *setpt)`  
*Sets new setpoint (without using config file). Write setpoint to FPGA. Before writing to FPGA the setpoint is divided by scaling parameter of 1st BPM. Scaling parameters are read from FPGA.*
- `int getSetpoint2 (char sel, double *setpt)`  
*Gets setpoint in [um]. Read setpoint from FPGA. After reading from FPGA the setpoint is multiplied by scaling parameter. Scaling parameters are read from CONFIG\_FILE.*
- `int IIRcontrollerSetParam (char ctrlr_sel, char param_is_a, char param_sel, double param)`  
*Sets IIR controller parameter.*
- `int IIRcontrollerCommitParams (char ctrlr_sel)`  
*Commits IIR controller parameters.*
- `int IIRcontrollerGetParam (char ctrlr_sel, char param_is_a, char param_sel, double *param)`  
*Reads IIR controller parameter.*
- `int funcGenConfig (float freq, float ampl[4], float offset[4])`  
*Configures function generator.*
- `int funcGenEnable (char enable)`  
*Enables or disables function generator.*
- `int getSFPInfo (int selector, uint16_t *id, uint32_t *timestamp, uint32_t *fw_ver, uint32_t *ser_nr)`  
*Gets SFP info.*
- `int getSystemVersion (uint32_t *hwVer, uint32_t *swVer, uint32_t *ctrlVer)`  
*Gets System HW, SW and PID versions.*

## Variables

- `double * in`
- `fftw_complex * out`
- `fftw_plan p`
- `int fd_DMA`
- `int fd_Mbox`
- `int fd_Lock`
- `int SAMP_FREQ = 1000`
- `int loginLevel = 0`
- `char LIB_BEST_VERSION [] = VERSION_STRING(FROM_DEFS_VERSION)`

### 7.1.1 Macro Definition Documentation

### 7.1.1.1 debug\_print

```
#define debug_print(  
    fmt,  
    ... ) do { if (LIB_BEST_DEBUG) fprintf(stderr, "[libbest] " fmt, __VA_ARGS__);  
} while (0)
```

### 7.1.1.2 LIB\_BEST\_DEBUG

```
#define LIB_BEST_DEBUG 0
```

### 7.1.1.3 VERSION\_STRING

```
#define VERSION_STRING(  
    X ) VERSION_STRING_HELPER(X)
```

### 7.1.1.4 VERSION\_STRING\_HELPER

```
#define VERSION_STRING_HELPER(  
    X ) #X
```

## 7.1.2 Function Documentation

### 7.1.2.1 \_\_attribute\_\_() [1/2]

```
void __attribute__ (  
    (constructor) )
```

### 7.1.2.2 \_\_attribute\_\_() [2/2]

```
void __attribute__ (  
    (destructor) )
```

### 7.1.2.3 getBestLock()

```
int getBestLock (  
    int fd )  
!!! All sets should check loginLevel
```

### 7.1.2.4 lock\_sigaction()

```
void lock_sigaction (  
    int sig,  
    siginfo_t * siginfo,  
    void * context )
```

### 7.1.2.5 releaseBestLock()

```
int releaseBestLock (  
    int fd )
```

### 7.1.3 Variable Documentation

#### 7.1.3.1 fd\_DMA

```
int fd_DMA
```

#### 7.1.3.2 fd\_Lock

```
int fd_Lock
```

#### 7.1.3.3 fd\_Mbox

```
int fd_Mbox
```

#### 7.1.3.4 in

```
double* in
```

#### 7.1.3.5 LIB\_BEST\_VERSION

```
char LIB_BEST_VERSION[ ] = VERSION_STRING(FROM_DEFS_VERSION)
```

#### 7.1.3.6 loginLevel

```
int loginLevel = 0
```

#### 7.1.3.7 out

```
fftw_complex* out
```

#### 7.1.3.8 p

```
fftw_plan p
```

#### 7.1.3.9 SAMP\_FREQ

```
int SAMP_FREQ = 1000
```

## 7.2 best\_c\_interface.h File Reference

```
#include <stdint.h>
```

### Macros

- #define BEST\_FILE\_MAILBOX "/dev/best\_mailbox"
- #define BEST\_FILE\_DISP\_DMA "/dev/best\_dma\_displ"
- #define LOCK\_FILE "/var/lock/best\_lock"
- #define BEST\_FILE\_CONFIG\_INI "/opt/CAENels/BEST/config.ini"

- #define FFT\_LEN (2048)
- #define LEN\_BBF (32)

## Enumerations

- enum `best_pid_select` { `BEST_PID_SELECT_X` = 0, `BEST_PID_SELECT_Y` = 1, `BEST_PID_SELECT_I0` = 2 }

## Functions

- struct `__attribute__((packed)) best_buffer_line`
- int `getBuffer` (struct `best_buffer_line`[], int length)
 

*Gets arrays of "samples" from display DMA.*
- int `getPosArray_sel` (int selector, double \*posX, double \*posY, double \*I0, unsigned int length)
 

*Gets arrays of positions and intensity from display DMA.*
- int `getPosScalar_sel` (int selector, double \*posX, double \*posY, double \*I0, unsigned int length)
 

*Gets arrays of positions and intensity from display DMA.*
- int `getPosArray` (double \*posX, double \*posY, double \*I0, unsigned int length)
 

*Same as `getPosArray_sel`, but only for 1st TetrAMM.*
- int `getVoltageArray` (double(\*voltages)[4], unsigned int length)
 

*Gets arrays of voltages from display DMA.*
- int `getVoltageArray_sel` (int sel, double(\*voltages)[4], unsigned int length)
 

*Gets arrays of voltages one of X PreDAC (the PreDAC is only 1 for the time being)*
- int `getCurrentArray` (double(\*currents)[4], unsigned int length)
 

*Gets arrays of currents from display DMA.*
- int `getCurrentArray_sel` (int sel, double(\*currents)[4], unsigned int length)
 

*Gets arrays of currents one of two TetrAMMs.*
- int `getFFT` (unsigned int selector, double \*outM, double \*outF, int removeDC)
 

*Calculates FFT.*
- int `getPosStats` (int selector, double \*meanX, double \*stdX, double \*stdX\_filt, double \*meanY, double \*stdY, double \*stdY\_filt, double \*meanI0, double \*stdI0, double \*stdI0\_filt)
 

*Get position statistics.*
- int `setDisplayFreq` (int freq, uint8\_t use\_filter)
 

*Sets display frequency.*
- int `getDisplayFreq` (int \*freq, uint8\_t \*use\_filter)
 

*Gets display frequency.*
- struct `__attribute__((__packed__)) best_stats`
- int `getBPMStats` (`best_stats` \*stats)
 

*Gets BPM Statistics.*
- int `getBPMscaling_sel` (int selector, double \*scaleX, double \*scaleY)
 

*Gets BPM scaling parameters for position X and Y in [um]. Read values from CONFIG\_FILE.*
- int `getBPMscaling` (double \*scaleX, double \*scaleY)
 

*Gets BPM scaling parameters of 1st BPM for position X and Y in [um]. Read values from CONFIG\_FILE.*
- int `getROCanROI` (double \*roiX, double \*roiY, double \*roi0min, double \*roi0max, double \*roc)
 

*Gets RoiX, RoiY, Roi0max, Roi0min and Roc of 1st BPM. Read values from CONFIG\_FILE.*
- int `getROCenable` (bool \*rocX, bool \*rocY, bool \*rocl0, bool \*rocX2, bool \*rocY2, bool \*rocl02)
 

*Gets enables for different ROC checks of 1st and 2nd BPMs. Read values from FPGA.*
- int `setROCenable` (bool rocX, bool rocY, bool rocl0, bool rocX2, bool rocY2, bool rocl02)
 

*Sets enables for different ROC checks of 1st and 2nd BPMs. Writes values to FPGA.*
- int `setOffsetSingle` (enum `best_pid_select` pid\_sel, double offset)
 

*Sets output offset in [V] (output=this+PID value, updated even when PID is not running). Write offset to FPGA.*
- int `getOffsetSingle` (enum `best_pid_select` pid\_sel, double \*offset)

- Gets output offset in [V]. Read offset from FPGA.
- int **setOffset** (double offsetX, double offsetY, double offsetI0)
  - Sets output offsets in [V] (output=this+PID value, updated even when PID is not running).*
- int **setWindAvg** (enum **best\_pid\_select** pid\_sel, uint32\_t nr\_samp)
  - Sets number of samples for window averaging.*
- int **getWindAvg** (enum **best\_pid\_select** pid\_sel, uint32\_t \*nr\_samp)
  - Gets number of samples for window averaging. Read number of samples from FPGA.*
- int **setPIDparams** (char pid\_sel, double Kp, double Ki, double Kd, double emin, double lmax, double Omin, double Omax, double Ogain)
  - Sets PID parameters (legacy).*
- int **setPIDparamSingle** (char pid\_sel, int pid\_par, double param\_value)
  - Sets single PID parameter of specific PID. Write PID parameter to FPGA.*
- int **getPIDparamSingle** (char pid\_sel, int pid\_par, double \*param\_value)
  - Gets PID parameters. Read PID parameters from FPGA.*
- int **setPIDoutSel** (uint8\_t posX, uint8\_t posY, uint8\_t I0)
  - Sets PID output cross switch. Selects which PID is assigned to which PreDAC output channel. Write PID output cross switch to FPGA.*
- int **getPIDoutSel** (uint8\_t \*posX, uint8\_t \*posY, uint8\_t \*I0)
  - Gets PID output cross switch. Read PID output cross switch from FPGA.*
- int **setCrossBar** (uint8\_t sel, uint8\_t ch0, uint8\_t ch1, uint8\_t ch2, uint8\_t ch3)
  - Sets input cross switch.*
- int **setBPMorient** (uint8\_t sel, uint8\_t is90deg)
  - Sets BPM orientation for diffOverSum.*
- int **setOutputMux** (uint8\_t sel)
  - Sets value of output mux.*
- int **getOutputMux** (uint8\_t \*sel)
  - Gets value of output mux.*
- int **setBPMscale** (double scaleX, double scaleY, int bpm)
  - Sets BPM scaling parameters for position X and Y in [um]. Write BPM scaling parameters to FPGA.*
- int **getBPMscale** (int selector, double \*scaleX, double \*scaleY)
  - Gets BPM scaling parameters for position X and Y in [um]. Read BPM scaling parameters from FPGA.*
- int **setBPMoffset** (double offsetX, double offsetY, int bpm)
  - Sets BPM offset in [um] ONLY FOR TETRAMMs. Write BPM offsets to FPGA.*
- int **setBPMoffsetEnbox** (double offsetX, double offsetY, int bpm, int encType)
  - Sets BPM offset in [um] ONLY FOR Enboxes. Write BPM offsets to FPGA. (>= 1.8-16)*
- int **setBPMdevice** (uint8\_t sel, uint8\_t isEnbox)
  - Sets BPM device for diffOverSum. Write to FPGA. (>= 1.8-18).*
- int **setEnboxType** (uint8\_t sel, uint8\_t encType)
  - Sets the EnBOX type (relative or absolute) . Write to FPGA. (>= 1.8-18).*
- int **setEnboxSumDiff** (uint8\_t sel, uint8\_t isSumDiff)
  - Set diffOverSum calculations (for EnBOX). 0:Normal the positions are calculated from enc1 and enc2. 1:SumDiff the positions are calculated from enc1+enc2 and enc1-enc2. Write to FPGA. (>= 1.8-18).*
- int **getBPMoffset** (double \*offsetX, double \*offsetY, int bpm)
  - Gets BPM offsets in [um] ONLY FOR TETRAMMs. Read BPM offsets from FPGA.*
- int **getBPMoffsetEnbox** (double \*offsetX, double \*offsetY, int bpm, int encType)
  - Gets BPM offsets in [um] ONLY FOR Enboxes. Read BPM offsets from FPGA.*
- int **getBPMorient** (uint8\_t sel, uint8\_t \*is90deg)
  - Sets BPM orientation for diffOverSum. Read BPM orientation from FPGA.*
- int **getBPMselector** (int \*posX, int \*posY, int \*I0)
  - Gets BPM sources. Read BPM sources from FPGA.*
- int **setBPMselector** (int posX, int posY, int I0)
  - Sets BPM selector. Write BPM selector to FPGA.*

- int **getCrossBar** (uint8\_t sel, uint8\_t \*ch0, uint8\_t \*ch1, uint8\_t \*ch2, uint8\_t \*ch3)  
*Gets input cross switch. Read input cross switch from FPGA.*
- int **setROCLimits** (int sel, double min, double max)  
*Sets ROC min and max values. ROC should be in [um] for positions and in [A] for intensity. Write min and max ROC levels of specific axis to FPGA.*
- int **getROCLimits** (int sel, double \*min, double \*max)  
*Gets ROC. ROC values are in [um] for positions and in [A] for intensity. Read min and max ROC levels of specific axis from FPGA.*
- int **getLock** (const char \*usr, const char \*pass)  
*Acquires lock on /var/lock/best\_lock.*
- int **getLockStatus** ()  
*Gets login level, which can be changed with call on **getLock()**.*
- unsigned int **getPIDstatus** (void)  
*Gets PID status.*
- int **setFBenable** (int enable)  
*Enables or disables feedback.*
- int **setCtrlReset** ()  
*Resets internal states of controller.*
- int **setPIDconf** (char conf)  
*Sets PID controller configuration. Write PID controller configuration to FPGA.*
- int **getPIDconf** (char \*conf)  
*Sets PID controller configuration. Read PID configuration from FPGA.*
- int **setSetpoint** (char sel, double setpt)  
*Sets new setpoint in [um]. Write setpoint to FPGA. Before writing to FPGA the setpoint is divided by scaling parameter of 1st BPM. Scaling parameters are read from CONFIG\_FILE.*
- int **getSetpoint** (char sel, double \*setpt)  
*Gets setpoint in [um]. Read setpoint from FPGA. After reading from FPGA the setpoint is multiplied by scaling parameter. Scaling parameters are read from CONFIG\_FILE.*
- int **setSetpoint2** (char sel, double setpt)  
*Sets new setpoint (without using config file). Write setpoint to FPGA. Before writing to FPGA the setpoint is divided by scaling parameter of 1st BPM. Scaling parameters are read from FPGA.*
- int **getSetpoint2** (char sel, double \*setpt)  
*Gets new setpoint (without using config file). Read setpoint from FPGA. After reading from FPGA the setpoint is multiplied by scaling parameter. Scaling parameters are read from FPGA.*
- int **IIRControllerSetParam** (char ctrlr\_sel, char param\_is\_a, char param\_sel, double param)  
*Sets IIR controller parameter.*
- int **IIRControllerCommitParams** (char ctrlr\_sel)  
*Commits IIR controller parameters.*
- int **IIRControllerGetParam** (char ctrlr\_sel, char param\_is\_a, char param\_sel, double \*param)  
*Reads IIR controller parameter.*
- int **funcGenConfig** (float freq, float ampl[4], float offset[4])  
*Configures function generator.*
- int **funcGenEnable** (char enable)  
*Enables or disables function generator.*
- int **tetrammSetRange** (int range, unsigned char selector)  
*Sets front-end range.*
- int **tetrammHVSetVoltage** (float voltage, unsigned char sel)  
*Sets TetrAMM HV voltage.*
- int **tetrammHVenable** (int enable, unsigned char sel)  
*Enables or disables TetrAMM HV module.*
- unsigned int **getTetrammsNumber** ()

- int [getTetramms](#) (int \*pos\_A, int \*pos\_B)  
*Gets number of TetrAMMs.*  
*Gets TetrAMMs. (>= 1.8-14).*
- int [enboxEncoderSelect](#) (int enc\_type, unsigned char selector)
- unsigned int [getEnboxesNumber](#) ()  
*Gets number of Enbox. (>= 1.8-15).*
- int [getEnboxes](#) (int \*pos\_A, int \*pos\_B)  
*Gets Enbox. (>= 1.8-15).*
- int [getSFPInfo](#) (int selector, uint16\_t \*id, uint32\_t \*timestamp, uint32\_t \*fw\_ver, uint32\_t \*ser\_nr)  
*Gets SFP info.*
- int [getSystemVersion](#) (uint32\_t \*hwVer, uint32\_t \*swVer, uint32\_t \*ctrlVer)  
*Gets System HW, SW and PID versions.*

## Variables

- char [LIB\\_BEST\\_VERSION](#) []

### 7.2.1 Macro Definition Documentation

#### 7.2.1.1 [BEST\\_FILE\\_CONFIG\\_INI](#)

```
#define BEST_FILE_CONFIG_INI "/opt/CAENels/BEST/config.ini"
```

#### 7.2.1.2 [BEST\\_FILE\\_DISP\\_DMA](#)

```
#define BEST_FILE_DISP_DMA "/dev/best_dma_disp1"
```

#### 7.2.1.3 [BEST\\_FILE\\_MAILBOX](#)

```
#define BEST_FILE_MAILBOX "/dev/best_mailbox"
```

#### 7.2.1.4 [FFT\\_LEN](#)

```
#define FFT_LEN (2048)
```

#### 7.2.1.5 [LEN\\_BBF](#)

```
#define LEN_BBF (32)
```

#### 7.2.1.6 [LOCK\\_FILE](#)

```
#define LOCK_FILE "/var/lock/best_lock"
```

### 7.2.2 Variable Documentation

#### 7.2.2.1 [LIB\\_BEST\\_VERSION](#)

```
char LIB_BEST_VERSION[ ]
```

## 7.3 best\_c\_interface\_conf.cpp File Reference

```
#include "best_c_interface.h"
#include <stdlib.h>
#include <time.h>
#include <stdint.h>
#include <string.h>
#include <errno.h>
#include <unistd.h>
#include <fcntl.h>
#include <sys/stat.h>
#include <sys/ioctl.h>
#include "pcie_driver/BEST_PCIE.h"
#include "pcie_mailbox/mailbox_comm_defs.h"
#include <iostream>
#include "inih/cpp/INIReader.h"
#include <complex>
#include <fftw3.h>
#include <signal.h>
```

## Functions

- `unsigned int getPIDstatus (void)`  
*Gets PID status.*
- `int getPosStats (int selector, double *meanX, double *stdX, double *stdX_filt, double *meanY, double *stdY, double *stdY_filt, double *meanI0, double *stdI0, double *stdI0_filt)`  
*Get position statistics.*
- `int getBPMscaling_sel (int selector, double *scaleX, double *scaleY)`  
*Gets BPM scaling parameters for position X and Y in [um]. Read values from CONFIG\_FILE.*
- `int getBPMscaling (double *scaleX, double *scaleY)`  
*Gets BPM scaling parameters of 1st BPM for position X and Y in [um]. Read values from CONFIG\_FILE.*
- `int getBPMscale (int selector, double *scaleX, double *scaleY)`  
*Gets BPM scaling parameters for position X and Y in [um]. Read BPM scaling parameters from FPGA.*
- `int getROCanROI (double *roiX, double *roiY, double *roi0min, double *roi0max, double *roc)`  
*Gets RoiX, RoiY, Roi0max, Roi0min and Roc of 1st BPM. Read values from CONFIG\_FILE.*
- `int getROCeable (bool *rocX, bool *rocY, bool *rocl0, bool *rocX2, bool *rocY2, bool *rocl02)`  
*Gets enables for different ROC checks of 1st nad 2nd BPMs. Read values from FPGA.*
- `int setROCeable (bool rocX, bool rocY, bool rocl0, bool rocX2, bool rocY2, bool rocl02)`  
*Sets enables for different ROC checks of 1st nad 2nd BPMs. Writes values to FPGA.*
- `int setOffsetSingle (enum best_pid_select pid_sel, double offset)`  
*Sets output offset in [V] (output=this+PID value, updated even when PID is not running). Write offset to FPGA.*
- `int getOffsetSingle (enum best_pid_select pid_sel, double *offset)`  
*Gets output offset in [V]. Read offset from FPGA.*
- `int setOffset (double offsetX, double offsetY, double offsetI0)`  
*Sets output offsets in [V] (output=this+PID value, updated even when PID is not running).*
- `int setWindAvg (enum best_pid_select pid_sel, uint32_t nr_samp)`  
*Sets number of samples for window averaging.*
- `int getWindAvg (enum best_pid_select pid_sel, uint32_t *nr_samp)`  
*Gets number of samples for window averaging. Read number of samples from FPGA.*
- `int setPIDparams (char pid_sel, double Kp, double Ki, double Kd, double emin, double lmax, double Omin, double Omax, double Ogain)`  
*Sets PID parameters (legacy).*
- `int setPIDparamSingle (char pid_sel, int pid_par, double param_value)`

- int **getPIDparamSingle** (char pid\_sel, int pid\_par, double \*param\_value)
 

*Sets single PID parameter of specific PID. Write PID parameter to FPGA.*
- int **setPIDoutSel** (uint8\_t posX, uint8\_t posY, uint8\_t l0)
 

*Gets PID parameters. Read PID parameters from FPGA.*
- int **setPIDoutSel** (uint8\_t \*posX, uint8\_t \*posY, uint8\_t \*l0)
 

*Sets PID output cross switch. Selects which PID is assigned to which PreDAC output channel. Write PID output cross switch to FPGA.*
- int **getPIDoutSel** (uint8\_t \*posX, uint8\_t \*posY, uint8\_t \*l0)
 

*Gets PID output cross switch. Read PID output cross switch from FPGA.*
- int **setCrossBar** (uint8\_t sel, uint8\_t ch0, uint8\_t ch1, uint8\_t ch2, uint8\_t ch3)
 

*Sets input cross switch.*
- int **setBPMorient** (uint8\_t sel, uint8\_t is90deg)
 

*Sets BPM orientation for diffOverSum.*
- int **getBPMorient** (uint8\_t sel, uint8\_t \*is90deg)
 

*Sets BPM orientation for diffOverSum. Read BPM orientation from FPGA.*
- int **setOutputMux** (uint8\_t sel)
 

*Sets value of output mux.*
- int **getOutputMux** (uint8\_t \*sel)
 

*Gets value of output mux.*
- int **setBPMscale** (double scaleX, double scaleY, int bpm)
 

*Sets BPM scaling parameters for position X and Y in [um]. Write BPM scaling parameters to FPGA.*
- int **setBPMoffset** (double offsetX, double offsetY, int bpm)
 

*Sets BPM offset in [um] ONLY FOR TETRAMMs. Write BPM offsets to FPGA.*
- int **setBPMoffsetEnbox** (double offsetX, double offsetY, int bpm, int encType)
 

*Sets BPM offset in [um] ONLY FOR Enboxes. Write BPM offsets to FPGA. (>= 1.8-16)*
- int **setBPMdevice** (uint8\_t sel, uint8\_t isEnbox)
 

*Sets BPM device for diffOverSum. Write to FPGA. (>= 1.8-18).*
- int **setEnboxType** (uint8\_t sel, uint8\_t encType)
 

*Sets the EnBOX type (relative or absolute) . Write to FPGA. (>= 1.8-18).*
- int **setEnboxSumDiff** (uint8\_t sel, uint8\_t isSumDiff)
 

*Set diffOverSum calculations (for EnBOX). 0:Normal the positions are calculated from enc1 and enc2. 1:SumDiff the positions are calculated from enc1+enc2 and enc1-enc2. Write to FPGA. (>= 1.8-18).*
- int **getBPMoffset** (double \*offsetX, double \*offsetY, int bpm)
 

*Gets BPM offsets in [um] ONLY FOR TETRAMMs. Read BPM offsets from FPGA.*
- int **getBPMoffsetEnbox** (double \*offsetX, double \*offsetY, int bpm, int encType)
 

*Gets BPM offsets in [um] ONLY FOR Enboxes. Read BPM offsets from FPGA.*
- int **getBPMselector** (int \*posX, int \*posY, int \*l0)
 

*Gets BPM sources. Read BPM sources from FPGA.*
- int **setBPMselector** (int posX, int posY, int l0)
 

*Sets BPM sources. Write BPM sources to FPGA.*
- int **getCrossBar** (uint8\_t sel, uint8\_t \*ch0, uint8\_t \*ch1, uint8\_t \*ch2, uint8\_t \*ch3)
 

*Gets input cross switch. Read input cross switch from FPGA.*
- int **setROCLimits** (int sel, double min, double max)
 

*Sets ROC min and max values. ROC should be in [um] for positions and in [A] for intensity. Write min and max ROC levels of specific axis to FPGA.*
- int **getROCLimits** (int sel, double \*min, double \*max)
 

*Gets ROC. ROC values are in [um] for positions and in [A] for intensity. Read min and max ROC levels of specific axis from FPGA.*

## Variables

- int `loginLevel`
- int `fd_Mbox`
- int `fd_DMA`
- double \* `in`
- `fftw_complex` \* `out`
- `fftw_plan` `p`
- int `SAMP_FREQ`

### 7.3.1 Variable Documentation

#### 7.3.1.1 `fd_DMA`

```
int fd_DMA
```

#### 7.3.1.2 `fd_Mbox`

```
int fd_Mbox
```

#### 7.3.1.3 `in`

```
double* in
```

#### 7.3.1.4 `loginLevel`

```
int loginLevel
```

#### 7.3.1.5 `out`

```
fftw_complex* out
```

#### 7.3.1.6 `p`

```
fftw_plan p
```

#### 7.3.1.7 `SAMP_FREQ`

```
int SAMP_FREQ
```

## 7.4 best\_c\_interface\_data.cpp File Reference

```
#include "best_c_interface.h"
#include <stdlib.h>
#include <time.h>
#include <stdint.h>
#include <string.h>
#include <errno.h>
#include <unistd.h>
#include <fcntl.h>
#include <sys/stat.h>
```

---

```
#include <sys/ioctl.h>
#include "pcie_driver/BEST_PCIE.h"
#include "pcie_mailbox/mailbox_comm_defs.h"
#include <iostream>
#include "inlh/cpp/INIReader.h"
#include <complex>
#include <fftw3.h>
#include <signal.h>
```

## Functions

- int `getBuffer` (struct `best_buffer_line` buf[], int length)
 

*Gets arrays of "samples" from display DMA.*
- int `getPosArray_sel` (int selector, double \*posX, double \*posY, double \*l0, unsigned int length)
 

*Gets arrays of postions and intensity from display DMA.*
- int `getPosScalar_sel` (int selector, double \*posX, double \*posY, double \*l0, unsigned int length)
 

*Gets arrays of postions and intensity from display DMA.*
- int `getPosArray` (double \*posX, double \*posY, double \*l0, unsigned int length)
 

*Same as getPosArray\_sel, but only for 1st TetrAMM.*
- int `getVoltageArray_sel` (int sel, double(\*voltages)[4], unsigned int length)
 

*Gets arrays of voltages one of X PreDAC (the PreDAC is only 1 for for the time beign)*
- int `getVoltageArray` (double(\*voltages)[4], unsigned int length)
 

*Gets arrays of voltages from display DMA.*
- int `getCurrentArray_sel` (int sel, double(\*currents)[4], unsigned int length)
 

*Gets arrays of currents one of two TetrAMMs.*
- int `getCurrentArray` (double(\*currents)[4], unsigned int length)
 

*Gets arrays of currents from display DMA.*
- double `fftw_abs` (fftw\_complex f)
- int `getFFT` (unsigned int selector, double \*outM, double \*outF, int removeDC)
 

*Calcualtes FFT.*
- int `setDisplayFreq` (int freq, uint8\_t use\_filter)
 

*Sets display frequency.*
- int `getDisplayFreq` (int \*freq, uint8\_t \*use\_filter)
 

*Gets display frequency.*
- int `getBPMStats` (`best_stats` \*stats)
 

*Gets BPM Statistics.*

## Variables

- int `loginLevel`
- int `fd_Mbox`
- int `fd_DMA`
- double \* `in`
- fftw\_complex \* `out`
- fftw\_plan `p`
- int `SAMP_FREQ`

### 7.4.1 Function Documentation

#### 7.4.1.1 fftw\_abs()

```
double fftw_abs (
    fftw_complex f )
```

### 7.4.2 Variable Documentation

#### 7.4.2.1 fd\_DMA

```
int fd_DMA
```

#### 7.4.2.2 fd\_Mbox

```
int fd_Mbox
```

#### 7.4.2.3 in

```
double* in
```

#### 7.4.2.4 loginLevel

```
int loginLevel
```

#### 7.4.2.5 out

```
fftw_complex* out
```

#### 7.4.2.6 p

```
fftw_plan p
```

#### 7.4.2.7 SAMP\_FREQ

```
int SAMP_FREQ
```

## 7.5 best\_c\_interface\_tetramm.cpp File Reference

```
#include "best_c_interface.h"
#include <stdlib.h>
#include <time.h>
#include <stdint.h>
#include <string.h>
#include <errno.h>
#include <unistd.h>
#include <fcntl.h>
#include <sys/stat.h>
#include <sys/ioctl.h>
#include "pcie_driver/BEST_PCIE.h"
#include "pcie_mailbox/mailbox_comm_defs.h"
#include <iostream>
#include "inih/cpp/INIReader.h"
```

---

```
#include <complex>
#include <fftw3.h>
#include <signal.h>
```

## Functions

- int [tetrammSetRange](#) (int range, unsigned char selector)  
*Sets front-end range.*
- int [tetrammHVSetVoltage](#) (float voltage, unsigned char sel)  
*Sets TetrAMM HV voltage.*
- int [tetrammHVenable](#) (int enable, unsigned char sel)  
*Enables or disables TetrAMM HV module.*
- unsigned int [getTetrammsNumber](#) ()  
*Gets number of TetrAMMs.*
- unsigned int [getEnboxesNumber](#) ()  
*Gets number of Enbox. (>= 1.8-15).*
- int [getTetramms](#) (int \*pos\_A, int \*pos\_B)  
*Gets TetrAMMs. (>= 1.8-14).*
- int [getEnboxes](#) (int \*pos\_A, int \*pos\_B)  
*Gets Enbox. (>= 1.8-15).*
- int [enboxEncoderSelect](#) (int enc\_type, unsigned char selector)

## Variables

- int [loginLevel](#)
- int [fd\\_Mbox](#)

### 7.5.1 Variable Documentation

#### 7.5.1.1 [fd\\_Mbox](#)

```
int fd_Mbox
```

#### 7.5.1.2 [loginLevel](#)

```
int loginLevel
```

## 7.6 README.md File Reference

# Index

\_\_attribute\_\_  
    best\_c\_interface.c, 55  
    Data functions, 16

Access control functions, 38  
    getLock, 38  
    getLockStatus, 38

best\_buffer\_line, 51  
best\_c\_interface.c, 53  
    \_\_attribute\_\_, 55  
    debug\_print, 54  
    fd\_DMA, 56  
    fd\_Lock, 56  
    fd\_Mbox, 56  
    getBestLock, 55  
    in, 56  
    LIB\_BEST\_DEBUG, 55  
    LIB\_BEST\_VERSION, 56  
    lock\_sigaction, 55  
    loginLevel, 56  
    out, 56  
    p, 56  
    releaseBestLock, 55  
    SAMP\_FREQ, 56  
    VERSION\_STRING, 55  
    VERSION\_STRING\_HELPER, 55

best\_c\_interface.h, 56  
    BEST\_FILE\_CONFIG\_INI, 60  
    BEST\_FILE\_DISP\_DMA, 60  
    BEST\_FILE\_MAILBOX, 60  
    FFT\_LEN, 60  
    LEN\_BBF, 60  
    LIB\_BEST\_VERSION, 60  
    LOCK\_FILE, 60

best\_c\_interface\_conf.cpp, 61  
    fd\_DMA, 63  
    fd\_Mbox, 63  
    in, 63  
    loginLevel, 63  
    out, 63  
    p, 63  
    SAMP\_FREQ, 63

best\_c\_interface\_data.cpp, 63  
    fd\_DMA, 65  
    fd\_Mbox, 65  
    fftw\_abs, 64  
    in, 65  
    loginLevel, 65  
    out, 65

    p, 65  
    SAMP\_FREQ, 65

best\_c\_interface\_tetramm.cpp, 65  
    fd\_Mbox, 66  
    loginLevel, 66

BEST\_FILE\_CONFIG\_INI  
    best\_c\_interface.h, 60

BEST\_FILE\_DISP\_DMA  
    best\_c\_interface.h, 60

BEST\_FILE\_MAILBOX  
    best\_c\_interface.h, 60

best\_pid\_select  
    Configuration functions, 22

BEST\_PID\_SELECT\_IO  
    Configuration functions, 22

BEST\_PID\_SELECT\_X  
    Configuration functions, 22

BEST\_PID\_SELECT\_Y  
    Configuration functions, 22

best\_stats, 51

Configuration functions, 21  
    best\_pid\_select, 22  
    BEST\_PID\_SELECT\_IO, 22  
    BEST\_PID\_SELECT\_X, 22  
    BEST\_PID\_SELECT\_Y, 22  
    getBPMoffset, 23  
    getBPMoffsetEnbox, 23  
    getBPMorient, 23  
    getBPMscale, 24  
    getBPMscaling, 24  
    getBPMscaling\_sel, 24  
    getBPMselector, 25  
    getCrossBar, 25  
    getOffsetSingle, 26  
    getOutputMux, 26  
    getPIDoutSel, 26  
    getPIDparamSingle, 27  
    getROCanROI, 27  
    getROCenable, 28  
    getROClimits, 28  
    getWindAvg, 28  
    setBPMdevice, 29  
    setBPMoffset, 29  
    setBPMoffsetEnbox, 30  
    setBPMorient, 30  
    setBPMscale, 30  
    setBPMselector, 32  
    setCrossBar, 32  
    setEnboxSumDiff, 33

setEnboxType, 33  
 setOffset, 33  
 setOffsetSingle, 34  
 setOutputMux, 34  
 setPIDoutSel, 35  
 setPIDparams, 35  
 setPIDparamSingle, 36  
 setROCenable, 36  
 setROClimits, 37  
 setWindAvg, 37

Data functions, 15  
 \_\_attribute\_\_, 16  
 getBPMStats, 16  
 getBuffer, 16  
 getCurrentArray, 16  
 getCurrentArray\_sel, 17  
 getDisplayFreq, 17  
 getFFT, 17  
 getPosArray, 18  
 getPosArray\_sel, 18  
 getPosScalar\_sel, 18  
 getPosStats, 19  
 getVoltageArray, 19  
 getVoltageArray\_sel, 19  
 setDisplayFreq, 20  
 debug\_print  
     best\_c\_interface.c, 54

EnBOX functions, 47  
     enboxEncoderSelect, 47  
     getEnboxes, 47  
     getEnboxesNumber, 47  
 enboxEncoderSelect  
     EnBOX functions, 47

fd\_DMA  
     best\_c\_interface.c, 56  
     best\_c\_interface\_conf.cpp, 63  
     best\_c\_interface\_data.cpp, 65  
 fd\_Lock  
     best\_c\_interface.c, 56  
 fd\_Mbox  
     best\_c\_interface.c, 56  
     best\_c\_interface\_conf.cpp, 63  
     best\_c\_interface\_data.cpp, 65  
     best\_c\_interface\_tetramm.cpp, 66  
 FFT\_LEN  
     best\_c\_interface.h, 60  
 fftw\_abs  
     best\_c\_interface\_data.cpp, 64  
 funcGenConfig  
     Function generator functions, 44  
 funcGenEnable  
     Function generator functions, 44  
 Function generator functions, 44  
     funcGenConfig, 44  
     funcGenEnable, 44

getBestLock  
     best\_c\_interface.c, 55  
 getBPMoffset  
     Configuration functions, 23  
 getBPMoffsetEnbox  
     Configuration functions, 23  
 getBPMorient  
     Configuration functions, 23  
 getBPMscale  
     Configuration functions, 24  
 getBPMscaling  
     Configuration functions, 24  
 getBPMscaling\_sel  
     Configuration functions, 24  
 getBPMselector  
     Configuration functions, 25  
 getBPMStats  
     Data functions, 16  
 getBuffer  
     Data functions, 16  
 getCrossBar  
     Configuration functions, 25  
 getCurrentArray  
     Data functions, 16  
 getCurrentArray\_sel  
     Data functions, 17  
 getDisplayFreq  
     Data functions, 17  
 getEnboxes  
     EnBOX functions, 47  
 getEnboxesNumber  
     EnBOX functions, 47  
 getFFT  
     Data functions, 17  
 getLock  
     Access control functions, 38  
 getLockStatus  
     Access control functions, 38  
 getOffsetSingle  
     Configuration functions, 26  
 getOutputMux  
     Configuration functions, 26  
 getPIDconf  
     PID control functions, 39  
 getPIDoutSel  
     Configuration functions, 26  
 getPIDparamSingle  
     Configuration functions, 27  
 getPIDstatus  
     PID control functions, 40  
 getPosArray  
     Data functions, 18  
 getPosArray\_sel  
     Data functions, 18  
 getPosScalar\_sel  
     Data functions, 18  
 getPosStats  
     Data functions, 19

getROCandROI  
    Configuration functions, 27

getROCenable  
    Configuration functions, 28

getROClimits  
    Configuration functions, 28

getSetpoint  
    PID control functions, 40

getSetpoint2  
    PID control functions, 40

getSFPinfo  
    Misc, 49

getSystemVersion  
    Misc, 49

getTetramms  
    TetRAMM functions, 45

getTetrammsNumber  
    TetRAMM functions, 45

getVoltageArray  
    Data functions, 19

getVoltageArray\_sel  
    Data functions, 19

getWindAvg  
    Configuration functions, 28

IIRcontrollerCommitParams  
    PID control functions, 41

IIRcontrollerGetParam  
    PID control functions, 41

IIRcontrollerSetParam  
    PID control functions, 41

in  
    best\_c\_interface.c, 56  
    best\_c\_interface\_conf.cpp, 63  
    best\_c\_interface\_data.cpp, 65

LEN\_BBF  
    best\_c\_interface.h, 60

LIB\_BEST\_DEBUG  
    best\_c\_interface.c, 55

LIB\_BEST\_VERSION  
    best\_c\_interface.c, 56  
    best\_c\_interface.h, 60

LOCK\_FILE  
    best\_c\_interface.h, 60

lock\_sigaction  
    best\_c\_interface.c, 55

loginLevel  
    best\_c\_interface.c, 56  
    best\_c\_interface\_conf.cpp, 63  
    best\_c\_interface\_data.cpp, 65  
    best\_c\_interface\_tetramm.cpp, 66

Misc, 49  
    getSFPinfo, 49  
    getSystemVersion, 49

out  
    best\_c\_interface.c, 56

best\_c\_interface\_conf.cpp, 63  
best\_c\_interface\_data.cpp, 65

p  
    best\_c\_interface.c, 56  
    best\_c\_interface\_conf.cpp, 63  
    best\_c\_interface\_data.cpp, 65

PID control functions, 39  
    getPIDconf, 39  
    getPIDstatus, 40  
    getSetpoint, 40  
    getSetpoint2, 40  
    IIRcontrollerCommitParams, 41  
    IIRcontrollerGetParam, 41  
    IIRcontrollerSetParam, 41  
    setCtrlReset, 42  
    setFBenable, 42  
    setPIDconf, 42  
    setSetpoint, 42  
    setSetpoint2, 43

README.md, 66

releaseBestLock  
    best\_c\_interface.c, 55

SAMP\_FREQ  
    best\_c\_interface.c, 56  
    best\_c\_interface\_conf.cpp, 63  
    best\_c\_interface\_data.cpp, 65

setBPMdevice  
    Configuration functions, 29

setBPMoffset  
    Configuration functions, 29

setBPMoffsetEnbox  
    Configuration functions, 30

setBPMorient  
    Configuration functions, 30

setBPMscale  
    Configuration functions, 30

setBPMselector  
    Configuration functions, 32

setCrossBar  
    Configuration functions, 32

setCtrlReset  
    PID control functions, 42

setDisplayFreq  
    Data functions, 20

setEnboxSumDiff  
    Configuration functions, 33

setEnboxType  
    Configuration functions, 33

setFBenable  
    PID control functions, 42

setOffset  
    Configuration functions, 33

setOffsetSingle  
    Configuration functions, 34

setOutputMux  
    Configuration functions, 34

setPIDconf  
    PID control functions, [42](#)  
setPIDoutSel  
    Configuration functions, [35](#)  
setPIDparams  
    Configuration functions, [35](#)  
setPIDparamSingle  
    Configuration functions, [36](#)  
setROCenable  
    Configuration functions, [36](#)  
setROClimits  
    Configuration functions, [37](#)  
setSetpoint  
    PID control functions, [42](#)  
setSetpoint2  
    PID control functions, [43](#)  
setWindAvg  
    Configuration functions, [37](#)

TetrAMM functions, [45](#)  
    getTetramms, [45](#)  
    getTetrammsNumber, [45](#)  
    tetrammHVenable, [45](#)  
    tetrammHVSetVoltage, [46](#)  
    tetrammSetRange, [46](#)  
tetrammHVenable  
    TetrAMM functions, [45](#)  
tetrammHVSetVoltage  
    TetrAMM functions, [46](#)  
tetrammSetRange  
    TetrAMM functions, [46](#)

VERSION\_STRING  
    best\_c\_interface.c, [55](#)  
VERSION\_STRING\_HELPER  
    best\_c\_interface.c, [55](#)